H0535. TMA/RECRA

RECRA ENVIRONMENTAL INC.

Chemical and Environmental Measurement Information

Recra LabNet Philadelphia
Analytical Report

Client : TNU-HANFORD B99-085
RFW# : 9909L126
SDG/SAF #: H0535/B99-085

W.O. #: 10985-001-001-9999-00
Date Received: 09-17-99

SEMIVOLATILE

One (1) water sample was collected on 09-15-99.

The sample and its associated QC samples were extracted on 09-21-99 and analyzed according to criteria set forth in Recra OPs based on SW 846 Method 8270B TCL Semivolatile target compounds on 10-04-99.

The following is a summary of the QC results accompanying the sample results and a description of any problems encountered during their analyses:

1. The cooler temperature upon receipt has been recorded on the chain-of-custody.

2. The required holding times for extraction and analysis were met.

3. Non-target compounds were detected in these samples.

4. These samples were spectrally searched for Butylated Hydroxytoluene; however, it was not identified in the samples.

5. All surrogate recoveries were within USEPA QC limits.

6. Two (2) of eleven (11) matrix spike recoveries were outside USEPA QC limits. A copy of the Sample Discrepancy Report (SDR) has been enclosed.

7. Two (2) of eleven (11) blank spike recoveries were outside USEPA QC limits. A copy of the Sample Discrepancy Report (SDR) has been enclosed.

for J. Michael Taylor
Vice President
Philadelphia Analytical Laboratory

10-29-99
Date

pef\gorup\data\bna\tnu09126.doc

001

GLOSSARY OF BNA DATA

DATA QUALIFIERS

U = Compound was analyzed for but not detected. The associated numerical value is the estimated sample quantitation limit which is included and corrected for dilution and percent moisture.

J = Indicates an estimated value. This flag is used under the following circumstances: 1) when estimating a concentration for tentatively identified compounds (TICs) where a 1:1 response is assumed; or 2) when the mass spectral data indicate the presence of a compound that meets the identification criteria but the result is less than the specified detection limit but greater than zero. For example, if the limit of detection is 10 ug/L and a concentration of 3 ug/L is calculated, it is reported as 3J.

B = This flag is used when the analyte is found in the associated blank as well as in the sample. It indicates possible/probable blank contamination. This flag is also used for a TIC as well as for a positively identified TCL compound.

E = Indicates that the compound was detected beyond the calibration range and was subsequently analyzed at a dilution.

D = Identifies all compounds identified in an analysis at a secondary dilution factor.

I = Interference.

NQ = Result qualitatively confirmed but not able to quantify.

A = Indicates that a TIC is a suspected aldol-condensation product.

N = Indicates presumptive evidence of a compound. This flag is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It is applied to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the N code is not used.

X = This flag is used for a TIC compound which is quantified relative to a response factor generated from a daily calibration standard (rather than quantified relative to the closest internal standard).

Y = Additional qualifiers used as required are explained in the case narrative.

GLOSSARY OF BNA DATA

ABBREVIATIONS

BS = Indicates blank spike in which reagent grade water is spiked with the CLP matrix spike solutions and carried through all the steps in the method. Spike recoveries are reported.

BSD = Indicates blank spike duplicate.

MS = Indicates matrix spike.

MSD = Indicates matrix spike duplicate.

DL = Suffix added to sample number to indicate that results are from a diluted analysis.

NA = Not Applicable.

DF = Dilution Factor.

NR = Not Required.

SP, Z = Indicates Spiked Compound.

Recra LabNet Philadelphia **Sample Discrepancy Report (SDR)** SDR #: 99MS081

Initiator: S Durbe RFW Batch: 9909L126 Parameter: BNA
Date: 10-5-99 Samples: MSD, BS Matrix: water
Client: TNU Hanford 699-085 Method: SW846/MCAWW/CLP/ Prep Batch: 99LE1150 cont

1. Reason for SDR

a. COC Discrepancy __ Tech Profile Error __ Client Request __ Sampler Error on C-O-C
__ Transcription Error __ Wrong Test Code __ Other ______________

b. General Discrepancy

__ Missing Sample/Extract __ Container Broken __ Wrong Sample Pulled __ Label ID's Illegible
__ Hold Time Exceeded __ Insufficient Sample __ Preservation Wrong __ Received Past Hold
__ Improper Bottle Type __ Not Amenable to Analysis

Note : Verified by [Log-In] or [Prep Group] (circle)...signature/date: ______________

c. QC Problem (Include all relevant specific results; attach data if necessary)

MSD + BS yielded very low recoveries for 4-nitrophenol + pentachlorophenol

2. Known or Probable Causes(s) possible problem with prep

3. Discussion and Proposed Action Other Description: Narrate

__ Re-log
__ Entire Batch
__ Following Samples: ______
__ Re-leach
__ Re-extract
__ Re-digest
__ Revise EDD
__ Change Test Code to ______
__ Place On/Take Off Hold (circle)

4. Project Manager Instructions...signature/date: [signature] 10/7/99

__ Concur with Proposed Action
__ Disagree with Proposed Action; See Instruction
__ Include in Case Narrative
__ Client Contacted:
Date/Person ______________
__ Add
__ Cancel

5. Final Action...signature/date: KTD 10/25/99 Other Explanation:

__ Verified re-[log][leach][extract][digest][analysis] (circle)
X Included in Case Narrative
__ Hard Copy COC Revised
__ Electronic COC Revised
__ EDD Corrections Completed

When Final Action has been recorded, forward original to QA Specialist for distribution and filing.

Route	Distribution of Completed SDR	Route	Distribution of Completed SDR
__	X Initiator	__	__ Metals: Doughty
__	X Lab Manager: M. Taylor	__	__ Inorganic: Perrone
__	X Project Mgr: Stone/Carey/Schrenkel/Johnson	__	__ GC/LC: Schnell
__	X Section Mgr: Wesson/Daniels	__	__ MS: LeMin/Taylor
__	X QA (file): Racioppi	__	__ Log-in: Toder
__	__ Data Management: Feldman	__	__ Admin: Soos
__	__ Sample Prep: Schnell/Doughty/Kauffman	__	__ Other: ______

L-WI-006/E-03/99

	Cust ID:	B0WCP8		B0WCP8		B0WCP8		SBLKDI		SBLKDI BS	
Sample Information	RFW#:	001		001 MS		001 MSD		99LE1150-MB1		99LE1150-MB1	
	Matrix:	WATER		WATER		WATER		WATER		WATER	
	D.F.:	1.00		1.00		1.00		1.00		1.00	
	Units:	UG/L		UG/L		UG/L		UG/L		UG/L	
Surrogate Recovery	Nitrobenzene-d5	98	%	91	%	91	%	83	%	85	%
	2-Fluorobiphenyl	79	%	77	%	75	%	77	%	67	%
	Terphenyl-d14	89	%	93	%	82	%	82	%	87	%
	Phenol-d5	82	%	83	%	79	%	81	%	69	%
	2-Fluorophenol	70	%	79	%	75	%	82	%	49	%
	2,4,6-Tribromophenol	73	%	93	%	68	%	69	%	56	%
Phenol		10	U	76	%	76	%	10	U	65	%
bis(2-Chloroethyl)ether		10	U	20	U	20	U	10	U	10	U
2-Chlorophenol		10	U	73	%	74	%	10	U	59	%
1,3-Dichlorobenzene		10	U	20	U	20	U	10	U	10	U
1,4-Dichlorobenzene		10	U	72	%	72	%	10	U	48	%
1,2-Dichlorobenzene		10	U	20	U	20	U	10	U	10	U
2-Methylphenol		10	U	20	U	20	U	10	U	10	U
2,2'-oxybis(1-Chloropropane)		10	U	20	U	20	U	10	U	10	U
4-Methylphenol		10	U	20	U	20	U	10	U	10	U
N-Nitroso-di-n-propylamine		10	U	98	%	87	%	10	U	71	%
Hexachloroethane		10	U	20	U	20	U	10	U	10	U
Nitrobenzene		10	U	20	U	20	U	10	U	10	U
Isophorone		10	U	20	U	20	U	10	U	10	U
2-Nitrophenol		10	U	20	U	20	U	10	U	10	U
2,4-Dimethylphenol		10	U	20	U	20	U	10	U	10	U
bis(2-Chloroethoxy)methane		10	U	20	U	20	U	10	U	10	U
2,4-Dichlorophenol		10	U	20	U	20	U	10	U	10	U
1,2,4-Trichlorobenzene		10	U	77	%	80	%	10	U	54	%
Naphthalene		10	U	20	U	20	U	10	U	10	U
4-Chloroaniline		10	U	20	U	20	U	10	U	10	U
Hexachlorobutadiene		10	U	20	U	20	U	10	U	10	U
4-Chloro-3-methylphenol		10	U	81	%	73	%	10	U	72	%
2-Methylnaphthalene		10	U	20	U	20	U	10	U	10	U
Hexachlorocyclopentadiene		10	U	20	U	20	U	10	U	10	U
2,4,6-Trichlorophenol		10	U	20	U	20	U	10	U	10	U
2,4,5-Trichlorophenol		25	U	50	U	50	U	25	U	25	U

*= Outside of EPA CLP QC limits.

Cust ID:	B0WCP8	B0WCP8	B0WCP8	SBLKDI	SBLKDI BS
RFW#:	001	001 MS	001 MSD	99LE1150-MB1	99LE1150-MB1
2-Chloronaphthalene	10 U	20 U	20 U	10 U	10 U
2-Nitroaniline	25 U	50 U	50 U	25 U	25 U
Dimethylphthalate	10 U	20 U	20 U	10 U	10 U
Acenaphthylene	10 U	20 U	20 U	10 U	10 U
2,6-Dinitrotoluene	10 U	20 U	20 U	10 U	10 U
3-Nitroaniline	25 U	50 U	50 U	25 U	25 U
Acenaphthene	10 U	86 %	85 %	10 U	73 %
2,4-Dinitrophenol	25 U	50 U	50 U	25 U	25 U
4-Nitrophenol	25 U	32 %	0 * %	25 U	8 * %
Dibenzofuran	10 U	20 U	20 U	10 U	10 U
2,4-Dinitrotoluene	10 U	99 * %	84 %	10 U	69 %
Diethylphthalate	10 U	20 U	20 U	10 U	10 U
4-Chlorophenyl-phenylether	10 U	20 U	20 U	10 U	10 U
Fluorene	10 U	20 U	20 U	10 U	10 U
4-Nitroaniline	25 U	50 U	50 U	25 U	25 U
4,6-Dinitro-2-methylphenol	25 U	50 U	50 U	25 U	25 U
N-Nitrosodiphenylamine (1)	10 U	20 U	20 U	10 U	10 U
4-Bromophenyl-phenylether	10 U	20 U	20 U	10 U	10 U
Hexachlorobenzene	10 U	20 U	20 U	10 U	10 U
Pentachlorophenol	25 U	73 %	17 %	25 U	7 * %
Phenanthrene	10 U	20 U	20 U	10 U	10 U
Anthracene	10 U	20 U	20 U	10 U	10 U
Carbazole	10 U	20 U	20 U	10 U	10 U
Di-n-butylphthalate	1 J	2 J	2 J	10 U	10 U
Fluoranthene	10 U	20 U	20 U	10 U	10 U
Pyrene	10 U	94 %	83 %	10 U	88 %
Butylbenzylphthalate	10 U	20 U	20 U	10 U	10 U
3,3'-Dichlorobenzidine	10 U	20 U	20 U	10 U	10 U
Benzo(a)anthracene	10 U	20 U	20 U	10 U	10 U
Chrysene	10 U	20 U	20 U	10 U	10 U
bis(2-Ethylhexyl)phthalate	10 U	5 J	20 U	10 U	3 J
Di-n-octyl phthalate	10 U	20 U	20 U	10 U	10 U
Benzo(b)fluoranthene	10 U	20 U	20 U	10 U	10 U
Benzo(k)fluoranthene	10 U	20 U	20 U	10 U	10 U
Benzo(a)pyrene	10 U	20 U	20 U	10 U	10 U
Indeno(1,2,3-cd)pyrene	10 U	20 U	20 U	10 U	10 U
Dibenz(a,h)anthracene	10 U	20 U	20 U	10 U	10 U
Benzo(g,h,i)perylene	10 U	20 U	20 U	10 U	10 U

(1) - Cannot be separated from Diphenylamine. *= Outside of EPA CLP QC limits.

1F

SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

CLIENT SAMPLE NO.

B0WCP8

Lab Name: Recra.LabNet Work Order: 10985001001

Client: TNU-HANFORD B99-085

Matrix: (soil/water) WATER Lab Sample ID: 9909L126-001

Sample wt/vol: 1000 (g/mL) ML Lab File ID: A100410

Level: (low/med) LOW Date Received: 09/17/99

% Moisture: ____ decanted: (Y/N)__ Date Extracted: 09/21/99

Concentrated Extract Volume: 1000(uL) Date Analyzed: 10/04/99

Injection Volume: 2.0(uL) Dilution Factor: 1.00

GPC Cleanup: (Y/N) N pH: 7.0

Number TICs found: 3 CONCENTRATION UNITS: (ug/L or ug/Kg) UG/L

	CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.		UNKNOWN	7.77	2	J
2.		UNKNOWN	7.94	3	J
3.		UNKNOWN	23.13	4	J

1F

SEMIVOLATILE ORGANICS ANALYSIS DATA SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

CLIENT SAMPLE NO.

SBLKDI

Lab Name: Recra.LabNet Work Order: 10985001001

Client: TNU-HANFORD B99-085

Matrix: (soil/water) WATER Lab Sample ID: 99LE1150-MB1

Sample wt/vol: 1000 (g/mL) ML Lab File ID: A100408

Level: (low/med) LOW Date Received: 09/21/99

% Moisture: _____ decanted: (Y/N)__ Date Extracted: 09/21/99

Concentrated Extract Volume: 1000(uL) Date Analyzed: 10/04/99

Injection Volume: 2.0(uL) Dilution Factor: 1.00

GPC Cleanup: (Y/N) N pH: 7.0

Number TICs found: 0 CONCENTRATION UNITS: (ug/L or ug/Kg) UG/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				

Recra LabNet - Lionville Laboratory
BNA ANALYTICAL DATA PACKAGE FOR
TNU-HANFORD B99-085

DATE RECEIVED: 09/17/99 RFW LOT # :9909L126

CLIENT ID	RFW #	MTX	PREP #	COLLECTION	EXTR/PREP	ANALYSIS
B0WCP8	001	W	99LE1150	09/15/99	09/21/99	10/04/99
B0WCP8	001 MS	W	99LE1150	09/15/99	09/21/99	10/04/99
B0WCP8	001 MSD	W	99LE1150	09/15/99	09/21/99	10/04/99
LAB QC:						
SBLKDI	MB1	W	99LE1150	N/A	09/21/99	10/04/99
SBLKDI	MB1 BS	W	99LE1150	N/A	09/21/99	10/04/99

RECRA LabNet Use Only

9909L126

A11

Custody Transfer Record/Lab Work Request

Page 1 of 1

FIELD PERSONNEL: COMPLETE ONLY SHADED AREAS

(8) perrone wet chem

RECRA LabNet

Client: Thu Hanford B99-085

Est. Final Proj. Sampling Date:

Project #: 10985-001-001-9999-00

Project Contact/Phone #:

RECRA Project Manager: OJ

QC: spec Del: std TAT: 30 day

Date Rec'd: 9-17-99 Date Due: 10/17/99

Account #:

	VOA	BNA	Pest/PCB	Herb					Metal	CN	Sulfide	IC Anions pH	NO2 NO3
Refrigerator #	1	6							6		6	6	6
#/Type Container Liquid	3V	2AG							3P		1P	1P	1P
#/Type Container Solid													
Volume Liquid	40ml	1L							1L		1L	1L	1L
Volume Solid													
Preservatives									HNO3		ZNAC NaOH	–	H2SO4
RECRA LabNet Use Only	0024H OGCSC	0025H 0024H	08/17/9?						Met①		1SFD	Ang①	1N3N2

MATRIX CODES: S - Soil, SE - Sediment, SO - Solid, SL - Sludge, W - Water, O - Oil, A - Air, DS - Drum Solids, DL - Drum Liquids, L - EP/TCLP Leachate, WI - Wipe, X - Other, F - Fish

Lab ID	Client ID/Description	MS	MSD	Matrix	Date Collected	Time Collected	VOA	BNA	Pest/PCB	Herb					Metal	CN	Sulfide	IC Anions pH	NO2 NO3
001	BOWCP8			W	9-15-99	0650	✓	✓							✓		✓	✓	✓
002	BOWCP9			↓	↓	0518	✓												

Special Instructions:

saf # B99-085

9/23/99 - 1NH3N added to 001 per client cc.

COMPOSITE WASTE

DATE/REVISIONS:

OGCSC = 1-propanol, Ethanol
Met① 2. = As, Ba, Cd, Cr, Pb, Se, Ag, Cu,
3. Ni, V, Zn, Be
Ang① 4. = 1CCL, 1CFL, 1CNO2, 1CNO3, 1CPO4,
5. 1CSO4, 1PH, 1NH3N
6. Run matrix QC

RECRA LabNet Use Only

Samples were:
1) Shipped ✓ or Hand Delivered
Airbill # *
2) Ambient or (Chilled)
3) Received in Good Condition (Y) or N
4) Labels Indicate Properly Preserved (Y) or N
5) Received Within Holding Times IC, pH autofail Y or (N)

COC Tape was:
1) Present on Outer Package (Y) or N
2) Unbroken on Outer Package (Y) or N
3) Present on Sample (Y) or N
4) Unbroken on Sample (Y) or N
COC Record Present Upon Sample Rec't (Y) or N
Cooler Temp. 4.1 °C

Relinquished by	Received by	Date	Time
Fed Ex	T Murray	9-17-99	1020

Relinquished by	Received by	Date	Time
	ORIGINAL REWRITTEN		

Discrepancies Between Samples Labels and COC Record? (Y) or N
NOTES: Sulfide bottle = 1L not 500mL as chain indicates

*4235795295(el)

CHAIN OF CUSTODY / REQUEST FOR ANALYSIS RECORD

FLUOR DANIEL FERNALD | P.O. BOX 538704, CINCINNATI, OH 45253-8704

REFERENCE DOCUMENT NO.: SPL-99-0918

ORIGINAL MEF NO.: NA

CORRESPONDING REFERENCE DOCUMENT NO.:

RELEASE NO.: TOCO 19758	PROJECT NO.: 20200-PSP-0005	FOR SAMPLE RELATED PROBLEMS ACS CONTACT / PHONE: Audrey Hannum 4943	CONTRACT PURCHASE ORDER / TASK ORDER NO.: 965800217-034
PROJECT NAME: Area 3A/4A Subsurface PreDesign		REQUIRED REPORT DATE / LAB TAT: 21 DAYS	SAMPLE SHIPMENT DATE: 9-15-99
PROJECT CONTACT / PHONE: Christine Messerly 4619		RECEIVING LAB NAME: RECRA	SAMPLE SHIPPER (Print): RON HOUSTON
CHARGE NO.: 5CJ31	LOT MARKING NO.: NA	RECEIVING LAB ADDRESS: LIONVILLE, PA, 19341	OFF-SITE LAB CONTACT: BOB CAREY
SAMPLING TEAM (Print) & GROUP NAME / PHONE: JOYCE GRACE/EM/SMMP/4848 / Keith Payne EM 3267 / John VANDINE / Bob Minger			
SAMPLING TEAM (Signature) & Badge No.): Joyce Grace 10781 / Keith Payne 8388 / John Vandine 5536 / Bob Minger 76951			

ITEM NO.	SAMPLE NUMBER: FACTS ID	SAMPLE NUMBER: CUSTOMER ID / SAMPLE POINT	SAMPLE MATRIX	COMP	GRAB	FILTERED	COLLECTION: DATE (1999)	COLLECTION: TIME	CONTAINER: TYPE	CONTAINER: VOL.	NO. CONTRS	PRESERVATION	ASL	ANALYSES REQUESTED (If more space is required, use the SPECIAL INSTRUCTIONS block.)	PDF	OFF-SITE
1	20036161712483-6B-L		S Soil		√	N	09/13	1006	G	60 ml	1	Cool 4°C	B	*	Y	Y
2	20036161313A4A-SUB-TB5		W water		√	N	09/13	0830	G	40	3	H_2SO_4 pH<2 Cool 4°C	B	*	Y	Y
3																
4																
5																
6																
7																
8																
9																
10																

No samples below this line JJG 09/13/99

SPECIAL INSTRUCTIONS: * TAL I = Total volatiles ** All Trip Blanks have bubbles

RELINQUISHED

ITEM / REASON	RELINQUISHED BY (Signature) / AFFILIATION	BADGE NO.	DATE	TIME
1, 2, Release to SPL	[signature] EM/SMMP	1 76951	9/14/99	1307
1, 2 TO SHIP	[signature] SPL	2 76719	9-15-99	1300
		3		
		4		
		5		

RECEIVED

RECEIVED BY (Signature) / AFFILIATION	BADGE NO.	DATE	TIME
Karen Heffernan/SPL	1 7706	9/14/99	1307
Jansen	2		
	3	9/16/99	0930
	4		
	5		

DISTRIBUTION OF COPIES: WHITE — ON-SITE - RELEASE FILE / OFF-SITE ANALYTICAL LAB - RETURN TO FEMP; YELLOW — ON-SITE - DISTRIBUTE AS NEEDED / OFF-SITE ANALYTICAL LAB - RETURN TO FEMP; BLUE — SAMPLING TECH / PROJECT FILE

FS-F-3361 (09-21-95)

SHADED AREAS ARE TO BE COMPLETED BY THE SAMPLE RECEIVING GROUP.

* IF "N", THEN COMPLETE APPLICABLE NON-CONFORMANCE PROCEDURE

RECRA
ENVIRONMENTAL
INC.

Chemical and Environmental Measurement Information

Recra LabNet Philadelphia
Analytical Report

Client: TNU HANFORD B99-085
RFW #: 9909L126
SDG/SAF#: H0535/B99-085

W.O. #: #: 10985-001-001-9999-00
Date Received: 09-17-99

GC SCAN

The set of samples consisted of two (2) water samples collected on 09-15-99.

The samples and their associated QC samples were prepared on 09-23-99 and analyzed by methodology based on EPA Method 8015B for Ethanol and Butanol on 09-27-99.

The following is a summary of the QC results accompanying these sample results and a description of any problems encountered during their analyses:

1. The samples were packaged and stored as specified in the method protocol; the cooler temperature upon receipt has been recorded on the chain-of-custody.

2. The required holding time for analysis was met.

3. All initial calibrations associated with this data set were within acceptance criteria.

4. All continuing calibration standards analyzed prior to the sample extracts were within acceptance criteria.

5. Surrogates were not used for this analysis.

6. The blank spike recovery was within advisory control limits of 50%-150%.

7. All matrix spike recoveries were within advisory control limits of 50%-150%.

Pat C

J. Michael Taylor
Vice President
Philadelphia Analytical Laboratory

10-15-99
Date

r:\share\lc\gcscan\09-126.doc

The results presented in this report relate only to the analytical testing and conditions of the samples at receipt and during storage. All pages of this report are integral parts of the analytical data. Therefore, this report should only be reproduced in its entirety of 6 pages.

001

GLOSSARY OF OGCSC DATA

DATA QUALIFIERS

U = Indicates that the compound was analyzed for but not detected. The minimum detection limit for the sample (not the method detection limit) is reported with the U (e.g., 10U).

J = Indicates an estimated value. This flag is used in cases where a target analyte is detected at a level less than the lower quantification level. If the limit of quantification is 10 ug/L and a concentration of 3 ug/L is calculated, it is reported as 3J.

B = This flag is used when the analyte is found in the associated blank as well as in the sample. It indicates possible/probable blank contamination.

E = Indicates that the compound was detected beyond the calibration range and was subsequently analyzed at a dilution.

I = Interference.

ABBREVIATIONS

BS = Indicates blank spike in which reagent grade water is spiked with the CLP matrix spiking solutions and carried through all the steps in the method. Spike recoveries are reported.

BSD = Indicates blank spike duplicate.

MS = Indicates matrix spike.

MSD = Indicates matrix spike duplicate.

DL = Indicates that recoveries were not obtained because the extract had to be diluted for analysis.

NA = Not Applicable.

DF = Dilution Factor.

NR = Not Required.

SP = Indicates spiked compound.

Recra LabNet - Lionville Laboratory

GC SCAN

Report Date: 10/04/99 12:43

RFW Batch Number: **9909L126** Client: **TNU-HANFORD B99-085** Work Order: **10985-001-001-9999-00** Page: 1

003

	Cust ID:	B0WCP8	B0WCP8	B0WCP8	B0WCP9	BLK	BLK BS
Sample Information	RFW#:	001	001 MS	001 MSD	002	99LLC142-MB1	99LLC142-MB1
	Matrix:	WATER	WATER	WATER	WATER	WATER	WATER
	D.F.:	1.00	1.00	1.00	1.00	1.00	1.00
	Units:	mg/L	mg/L	mg/L	mg/L	mg/L	mg/L
n-Propyl Alcohol		5.0 U	92 %	104 %	5.0 U	5.0 U	91 %
Ethanol		5.0 U	5.0 U	5.0 U	5.0 U	5.0 U	5.0 U

10/4/99

U= Analyzed, not detected. J= Present below detection limit. B= Present in blank. NR= Not requested. NS= Not spiked.
%= Percent recovery. D= Diluted out. I= Interference. NA= Not Applicable. *= Outside of Advisory limits.

Recra LabNet - Lionville Laboratory
GCSC ANALYTICAL DATA PACKAGE FOR
TNU-HANFORD B99-085

DATE RECEIVED: 09/17/99 RFW LOT # :9909L126

CLIENT ID	RFW #	MTX	PREP #	COLLECTION	EXTR/PREP	ANALYSIS
B0WCP8	001	W	99LLC142	09/15/99	09/23/99	09/27/99
B0WCP8	001 MS	W	99LLC142	09/15/99	09/23/99	09/27/99
B0WCP8	001 MSD	W	99LLC142	09/15/99	09/23/99	09/27/99
B0WCP9	002	W	99LLC142	09/15/99	09/23/99	09/27/99
LAB QC:						
BLK	MB1	W	99LLC142	N/A	09/23/99	09/27/99
BLK	MB1 BS	W	99LLC142	N/A	09/23/99	09/27/99

RECRA LabNet Use Only
9909L126

Custody Transfer Record/Lab Work Request

Page 1 of 1

A11

FIELD PERSONNEL: COMPLETE ONLY SHADED AREAS

(8) perrone wet chem

RECRA LabNet

Client: THU Hanford B99-085

Est. Final Proj. Sampling Date:

Project #: 10985-001-001-9999-00

Project Contact/Phone #:

RECRA Project Manager: OJ

QC: spec Del: std TAT: 30 day

Date Rec'd: 9-17-99 Date Due: 10/17/99

Account #:

	VOA	BNA	Pest/PCB	Herb				Metal	CN	Sulfide	IC Anions PH	NO2 NO3
Refrigerator #	1	6						6		6	6	6
#/Type Container – Liquid	3V	2AG						3P		1P	1P	1P
#/Type Container – Solid												
Volume – Liquid	40ml	1L						1L		1L	1L	1L
Volume – Solid										ZNAC		
Preservatives								HNO3		NaOH	–	H2SO4
ANALYSES REQUESTED →	ORGANIC							INORG				
RECRA LabNet Use Only	0024H OGCSC	00225H 00244H						Met①		1SFD	Ang①	1N3N2

MATRIX CODES:
S - Soil; SE - Sediment; SO - Solid; SL - Sludge; W - Water; O - Oil; A - Air; DS - Drum Solids; DL - Drum Liquids; L - EP/TCLP Leachate; WI - Wipe; X - Other; F - Fish

Lab ID	Client ID/Description	Matrix QC Chosen (✓) MS	MSD	Matrix	Date Collected	Time Collected	VOA	BNA	Pest/PCB	Herb	Metal	CN	Sulfide	IC Anions	NO2 NO3
001	BOWCP8			W	9-15-99	0650	✓	✓			✓		✓	✓	✓
002	BOWCP9			↓	↓	0518	✓								

Special Instructions:

saf # B99-085

9/23/99 - 1NH3N added to 001 per client cc.

COMPOSITE WASTE

DATE/REVISIONS:

1. OGCSC = 1-propanol, Ethanol
2. Met① = as, Ba, Cd, Cr, Pb, Se, Ag, Cu,
3. Ni, V, Zn, Be
4. Ang① = 1CCL, 1CFL, 1CNO2, 1CNO3, 1CPO4,
5. 1CSO4, 1PH, 1NH3N
6. Run matrix QC

RECRA LabNet Use Only

Samples were:
1) Shipped ✓ or Hand Delivered ___
Airbill # *
2) Ambient or (Chilled)
3) Received in Good Condition (Y) or N
4) Labels Indicate Properly Preserved (Y) or N
5) Received Within Holding Times Y or (N) — IC PH out of hold

COC Tape was:
1) Present on Outer Package (Y) or N
2) Unbroken on Outer Package (Y) or N
3) Present on Sample (Y) or N
4) Unbroken on Sample (Y) or N
COC Record Present Upon Sample Rec't (Y) or N
Cooler Temp. 4.1 °C

Relinquished by	Received by	Date	Time	Relinquished by	Received by	Date	Time
Fed Ex	TMurray	9-17-99	1020		ORIGINAL REWRITTEN		

Discrepancies Between Samples Labels and COC Record? (Y) or N

NOTES: Sulfide bottle 1L not 500mL as chain indicates

CHAIN OF CUSTODY / REQUEST FOR ANALYSIS RECORD

FLUOR DANIEL FERNALD | P.O. BOX 538704, CINCINNATI, OH 45253-8704

REFERENCE DOCUMENT NO.: SPL-99-0918

ORIGINAL MEF NO.: NA | CORRESPONDING REFERENCE DOCUMENT NO.: | PAGE 1 OF 1

RELEASE NO.: 10000 19758	PROJECT NO.: 20200-PSP-0005	FOR SAMPLE RELATED PROBLEMS ACS CONTACT / PHONE: Audrey Hannum 4943	CONTRACT PURCHASE ORDER / TASK ORDER NO.: 96S00217-034
PROJECT NAME: Area 3A/4A Subsurface PreDesign		REQUIRED REPORT DATE / LAB TAT: 21 DAYS	SAMPLE SHIPMENT DATE: 9-15-99
PROJECT CONTACT / PHONE: Christine Messerly 4619		RECEIVING LAB NAME: RECRA	SAMPLE SHIPPER (Print): RON HOUSTON
CHARGE NO.: 5CJ31	LOT MARKING NO.: NA	RECEIVING LAB ADDRESS: LIONVILLE, PA, 19341	OFF-SITE LAB CONTACT: ROB CAREY
SAMPLING TEAM (Print) & GROUP NAME / PHONE: JOYCE GRACE/EM/SMMP/4848 / Kathy Payne EM 3267 / JOHN VANDINE / Bob Minges			
SAMPLING TEAM (Signature) & Badge No.: Joyce Grace 10781 / Kathy Payne 9388		Jo Vandine 5536 / Bob Minges 76951	

ITEM NO.	SAMPLE NUMBER: FACTS ID	SAMPLE NUMBER: CUSTOMER ID / SAMPLE POINT	SAMPLE MATRIX	COMP	GRAB	FILTERED	COLLECTION: DATE	COLLECTION: TIME	CONTAINER: TYPE	CONTAINER: VOL.	NO. CONTRS	PRESERVATION	ASL	ANALYSES REQUESTED (If more space is required, use the SPECIAL INSTRUCTIONS block)	RDF	OFF-SITE
1	20036161	12483-6B-L	S Soil		√	N	09/13/1999	0006	G	60 ml	1	Cool 4°C	B	*	Y	Y
2	20036161	3A4A-SUB-TB5	W Water		√	N	09/13/1999	0830	G	40 ml	3	H_2SO_4 pH<2 Cool 4°C	B	*	Y	Y
3																
4																
5																
6																
7																
8																
9																
10																

No samples below this line

JG 09/13/99

SPECIAL INSTRUCTIONS: * TAL I = Total volatiles ** All Trip Blanks have bubbles

RELINQUISHED: ITEM / REASON	RELINQUISHED BY (Signature) / AFFILIATION	BADGE NO.	DATE	TIME	RECEIVED: RECEIVED BY (Signature) / AFFILIATION	BADGE NO.	DATE	TIME
1, 2, Release to SPL	Bob Minges SMMP	1 76951	9/14/99	1307	Karen Heffernan/SPL	1 7706	9/14/99	1307
1, 2 TO SHIP	SPL	2 76719	9-15-99	1300		2		
		3			Jansen	3	9/16/99	0930
		4				4		
		5				5		

DISTRIBUTION OF COPIES: WHITE — ON-SITE - RELEASE FILE / OFF-SITE ANALYTICAL LAB - RETURN TO FEMP; YELLOW — ON-SITE - DISTRIBUTE AS NEEDED / OFF-SITE ANALYTICAL LAB - RETURN TO FEMP; BLUE — SAMPLING TECH / PROJECT FILE

FS-F-3361 (09-21-95)

SHADED AREAS ARE TO BE COMPLETED BY THE SAMPLE RECEIVING GROUP.

* IF "N", THEN COMPLETE APPLICABLE NON-CONFORMANCE PROCEDURE

RECRA ENVIRONMENTAL INC.
Chemical and Environmental Measurement Information

Recra LabNet Philadelphia
Analytical Report

Client : TNU-HANFORD B99-085
RFW# : 9909L126
SDG/SAF #: H0535/B99-085

W.O. #: 10985-001-001-9999-00
Date Received: 09-17-99

GC/MS VOLATILE

Two (2) water samples were collected on 09-15-99.

The samples and their associated QC samples were analyzed according to criteria set forth in Recra OPs based on SW 846 Method 8260A for TCL Volatile target compounds on 09-27,28-99.

The following is a summary of the QC results accompanying these sample results and a description of any problems encountered during their analyses:

1. The cooler temperature upon receipt has been recorded on the chain-of-custody.
2. The required holding time for analysis was met.
3. A non-target compound was detected in method blank 99LVN315-MB1.
4. One (1) of twenty-one (21) surrogate recoveries were outside EPA QC limits. The surrogate recovery for 1,2-Dichloroethane-d4 was biased slightly high, however all spike recoveries were within limits, therefore no significant impact on the data.
5. All matrix spike recoveries were within EPA QC limits.
6. All blank spike recoveries were within EPA QC limits.
7. The method blank 99LVH374 contained the common laboratory contaminant Methylene Chloride and Acetone at levels less than the CRQL and and the target compound 2-Butanone at a level less than the CRQL. The method blank 99LVN315 contained the common laboratory contaminants Methylene Chloride and Acetone at levels less than 3x and 1x the CRQL, respectively and target compounds 2-Butanone, Chloromethane and Bromomethane at a level less than the CRQL.

for [signature]
J. Michael Taylor
Vice President
Philadelphia Analytical Laboratory

10-29-99
Date

som\group\data\ voa\tnu09126.doc

208 Welsh Pool Road • Lionville, PA 19341-1333 • (610) 280-3000 • Fax (610) 280-3041

GLOSSARY OF VOA DATA

DATA QUALIFIERS

U = Compound was analyzed for but not detected. The associated numerical value is the estimated sample quantitation limit which is included and corrected for dilution and percent moisture.

J = Indicates an estimated value. This flag is used under the following circumstances: 1) when estimating a concentration for tentatively identified compounds (TICs) where a 1:1 response is assumed; or 2) when the mass spectral data indicate the presence of a compound that meets the identification criteria but the result is less than the specified detection limit but greater than zero. For example, if the limit of detection is 10 ug/L and a concentration of 3 ug/L is calculated, it is reported as 3J.

B = This flag is used when the analyte is found in the associated blank as well as in the sample. It indicates possible/probable blank contamination. This flag is also used for a TIC as well as for a positively identified TCL compound.

E = Indicates that the compound was detected beyond the calibration range and was subsequently analyzed at a dilution.

D = Identifies all compounds identified in an analysis at a secondary dilution factor.

I = Interference.

NQ = Result qualitatively confirmed but not able to quantify.

N = Indicates presumptive evidence of a compound. This flag is only used for tentatively identified compounds (TICs), where the identification is based on a mass spectral library search. It is applied to all TIC results. For generic characterization of a TIC, such as chlorinated hydrocarbon, the N code is not used.

X = This flag is used for a TIC compound which is quantified relative to a response factor generated from a daily calibration standard (rather than quantified relative to the closest internal standard).

Y = Additional qualifiers used as required are explained in the case narrative.

GLOSSARY OF VOA DATA

ABBREVIATIONS

BS = Indicates blank spike in which reagent grade water is spiked with the CLP matrix spike solutions and carried through all the steps in the method. Spike recoveries are reported.

BSD = Indicates blank spike duplicate.

MS = Indicates matrix spike.

MSD = Indicates matrix spike duplicate.

DL = Suffix added to sample number to indicate that results are from a diluted analysis.

NA = Not Applicable.

DF = Dilution Factor.

NR = Not Required.

SP, Z = Indicates Spiked Compound.

mmz\10-94\gloss.voa

Recra LabNet - Lionville Laboratory

Volatiles by GC/MS, HSL List — Report Date: 10/18/99 06:30

RFW Batch Number: **9909L126** Client: **TNU-HANFORD B99-085** Work Order: 10985001001 Page: 1a

	Cust ID:	B0WCP8	B0WCP8	B0WCP8	B0WCP9	VBLKRC	VBLKSS
Sample Information	RFW#:	001	001 MS	001 MSD	002	99LVH374-MB1	99LVN315-MB1
	Matrix:	WATER	WATER	WATER	WATER	WATER	WATER
	D.F.:	1.00	1.00	1.00	1.00	1.00	1.00
	Units:	UG/L	UG/L	UG/L	UG/L	UG/L	UG/L
Surrogate Recovery	Toluene-d8	96 %	99 %	95 %	104 %	102 %	105 %
	Bromofluorobenzene	90 %	89 %	87 %	95 %	96 %	102 %
	1,2-Dichloroethane-d4	96 %	97 %	96 %	114 %	106 %	112 %
		fl	fl	fl	fl	fl	fl
Chloromethane		10 U	6 J	2 J	10 U	10 U	1 J
Bromomethane		10 U	10 U	10 U	10 U	10 U	3 J
Vinyl Chloride		10 U	10 U	10 U	10 U	10 U	10 U
Chloroethane		10 U	10 U	10 U	10 U	10 U	10 U
Methylene Chloride		2 JB	6 B	4 JB	2 JB	4 J	11
Acetone		10 U	10 U	10 U	10 U	2 J	5 J
Carbon Disulfide		5 U	5 U	5 U	5 U	5 U	5 U
1,1-Dichloroethene		5 U	93 %	93 %	5 U	5 U	5 U
1,1-Dichloroethane		5 U	5 U	5 U	5 U	5 U	5 U
1,2-Dichloroethene (total)		5 U	5 U	5 U	5 U	5 U	5 U
Chloroform		5 U	5 U	5 U	5 U	5 U	5 U
1,2-Dichloroethane		5 U	5 U	5 U	5 U	5 U	5 U
2-Butanone		10 U	10 U	10 U	10 U	2 J	2 J
1,1,1-Trichloroethane		5 U	5 U	5 U	5 U	5 U	5 U
Carbon Tetrachloride		5 U	5 U	5 U	5 U	5 U	5 U
Bromodichloromethane		5 U	5 U	5 U	5 U	5 U	5 U
1,2-Dichloropropane		5 U	5 U	5 U	5 U	5 U	5 U
cis-1,3-Dichloropropene		5 U	5 U	5 U	5 U	5 U	5 U
Trichloroethene		5 U	88 %	89 %	5 U	5 U	5 U
Dibromochloromethane		5 U	5 U	5 U	5 U	5 U	5 U
1,1,2-Trichloroethane		5 U	5 U	5 U	5 U	5 U	5 U
Benzene		5 U	92 %	93 %	5 U	5 U	5 U
Trans-1,3-Dichloropropene		5 U	5 U	5 U	5 U	5 U	5 U
Bromoform		5 U	5 U	5 U	5 U	5 U	5 U
4-Methyl-2-pentanone		10 U	10 U	10 U	10 U	10 U	10 U
2-Hexanone		10 U	10 U	10 U	10 U	10 U	10 U
Tetrachloroethene		5 U	5 U	5 U	5 U	5 U	5 U
1,1,2,2-Tetrachloroethane		5 U	5 U	5 U	5 U	5 U	5 U
Toluene		5 U	95 %	92 %	5 U	5 U	5 U

*= Outside of EPA CLP QC limits.

C-14

Cust ID:	B0WCP8	B0WCP8	B0WCP8	B0WCP9	VBLKRC	VBLKSS
RFW#:	001	001 MS	001 MSD	002	99LVH374-MB1	99LVN315-MB1
Chlorobenzene	5 U	90 %	90 %	5 U	5 U	5 U
Ethylbenzene	5 U	5 U	5 U	5 U	5 U	5 U
Styrene	5 U	5 U	5 U	5 U	5 U	5 U
Xylene (total)	5 U	5 U	5 U	5 U	5 U	5 U

*= Outside of EPA CLP QC limits.

Cust ID: **VBLKSS BS**

Sample Information
RFW#: **99LVN315-MB1**
Matrix: WATER
D.F.: 1.00
Units: UG/L

Surrogate Recovery	Toluene-d8	106	%
	Bromofluorobenzene	100	%
	1,2-Dichloroethane-d4	115 *	%

Chloromethane	1	JB
Bromomethane	10	U
Vinyl Chloride	10	U
Chloroethane	10	U
Methylene Chloride	12	B
Acetone	5	JB
Carbon Disulfide	5	U
1,1-Dichloroethene	97	%
1,1-Dichloroethane	5	U
1,2-Dichloroethene (total)	5	U
Chloroform	5	U
1,2-Dichloroethane	5	U
2-Butanone	2	JB
1,1,1-Trichloroethane	5	U
Carbon Tetrachloride	5	U
Bromodichloromethane	5	U
1,2-Dichloropropane	5	U
cis-1,3-Dichloropropene	5	U
Trichloroethene	104	%
Dibromochloromethane	5	U
1,1,2-Trichloroethane	5	U
Benzene	111	%
Trans-1,3-Dichloropropene	5	U
Bromoform	5	U
4-Methyl-2-pentanone	10	U
2-Hexanone	10	U
Tetrachloroethene	5	U
1,1,2,2-Tetrachloroethane	5	U
Toluene	113	%

*= Outside of EPA CLP QC limits.

Cust ID: **VBLKSS BS**

RFW#: **99LVN315-MB1**

Chlorobenzene	106	%
Ethylbenzene	5	U
Styrene	5	U
Xylene (total)	5	U

*= Outside of EPA CLP QC limits.

1E

VOLATILE ORGANICS ANALYSIS SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

B0WCP8

Lab Name: Recra.LabNet Contract: 10985001001

Lab Code: Recra Case No.: _____ SAS No.: _____ SDG No.: _____

Matrix: (soil/water) WATER Lab Sample ID: 9909L126-001

Sample wt/vol: 5.00 (g/mL) ML Lab File ID: h092735

Level: (low/med) LOW Date Received: 09/17/99

% Moisture: not dec. _____ Date Analyzed: 09/28/99

Column: (pack/cap) CAP Dilution Factor: 1.00

Number TICs found: 0 CONCENTRATION UNITS: (ug/L or ug/Kg) UG/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				

1E

VOLATILE ORGANICS ANALYSIS SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

B0WCP9

Lab Name: Recra.LabNet Contract: 10985001001

Lab Code: Recra Case No.: _____ SAS No.: _____ SDG No.: _____

Matrix: (soil/water) WATER Lab Sample ID: 9909L126-002

Sample wt/vol: 5.00 (g/mL) ML Lab File ID: n092836

Level: (low/med) LOW Date Received: 09/17/99

% Moisture: not dec. _____ Date Analyzed: 09/28/99

Column: (pack/cap) CAP Dilution Factor: 1.00

Number TICs found: 0 CONCENTRATION UNITS: (ug/L or ug/Kg) UG/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				

1E

VOLATILE ORGANICS ANALYSIS SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

VBLKRC

Lab Name: Recra.LabNet Contract: 10985001001

Lab Code: Recra Case No.: _____ SAS No.: _____ SDG No.: _____

Matrix: (soil/water) WATER Lab Sample ID: 99LVH374-MB1

Sample wt/vol: 5.00 (g/mL) ML Lab File ID: h092726

Level: (low/med) LOW Date Received: 09/27/99

% Moisture: not dec. _____ Date Analyzed: 09/27/99

Column: (pack/cap) CAP Dilution Factor: 1.00

Number TICs found: 0 CONCENTRATION UNITS:
(ug/L or ug/Kg) UG/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.				

1E

VOLATILE ORGANICS ANALYSIS SHEET
TENTATIVELY IDENTIFIED COMPOUNDS

EPA SAMPLE NO.

VBLKSS

Lab Name: Recra.LabNet Contract: 10985001001

Lab Code: Recra Case No.: _____ SAS No.: _____ SDG No.: _____

Matrix: (soil/water) WATER Lab Sample ID: 99LVN315-MB1

Sample wt/vol: 5.00 (g/mL) ML Lab File ID: n092821

Level: (low/med) LOW Date Received: 09/28/99

% Moisture: not dec. _____ Date Analyzed: 09/28/99

Column: (pack/cap) CAP Dilution Factor: 1.00

Number TICs found: 1 CONCENTRATION UNITS:
(ug/L or ug/Kg) UG/L

CAS NUMBER	COMPOUND NAME	RT	EST. CONC.	Q
1.	SILOXANE	9.020	6	J

FORM 1 VOA-TIC 1/87 Rev.

Recra LabNet - Lionville Laboratory
VOA ANALYTICAL DATA PACKAGE FOR
TNU-HANFORD B99-085

DATE RECEIVED: 09/17/99 RFW LOT # :9909L126

CLIENT ID	RFW #	MTX	PREP #	COLLECTION	EXTR/PREP	ANALYSIS
B0WCP8	001	W	99LVH374	09/15/99	N/A	09/28/99
B0WCP8	001 MS	W	99LVH374	09/15/99	N/A	09/28/99
B0WCP8	001 MSD	W	99LVH374	09/15/99	N/A	09/28/99
B0WCP9	002	W	99LVN315	09/15/99	N/A	09/28/99
LAB QC:						
VBLKRC	MB1	W	99LVH374	N/A	N/A	09/27/99
VBLKSS	MB1	W	99LVN315	N/A	N/A	09/28/99
VBLKSS	MB1 BS	W	99LVN315	N/A	N/A	09/28/99

RECRA LabNet Use Only
9909L126

A11

Custody Transfer Record/Lab Work Request

Page 1 of 1

RECRA LabNet

FIELD PERSONNEL: COMPLETE ONLY SHADED AREAS

Client: Thu Hanford B99-085
Est. Final Proj. Sampling Date:
Project #: 10985-001-001-9999-00
Project Contact/Phone #:
RECRA Project Manager: OJ
QC: spec Del: std TAT: 30 day
Date Rec'd: 9-17-99 Date Due: 10/17/99
Account #:

	VOA	BNA	Pest/PCB	Herb					Metal	CN	Sulfide	IC Anions	pH	NO2 NO3			
Refrigerator #	1	6							6		6	6		6			
#/Type Container – Liquid	3V	2AG							3P		1P	1P		1P			
#/Type Container – Solid																	
Volume – Liquid	40ml	1L							1L		1L	1L		1L			
Volume – Solid											ZNAC						
Preservatives									HNO3		NaOH	–		H2SO4			
RECRA LabNet Use Only	0024H OGCSC	0025H 0024H	0017H PCB						Met①		1SFD	Ang①		1N3N2			

MATRIX CODES: S - Soil; SE - Sediment; SO - Solid; SL - Sludge; W - Water; O - Oil; A - Air; DS - Drum Solids; DL - Drum Liquids; L - EP/TCLP Leachate; WI - Wipe; X - Other; F - Fish

Lab ID	Client ID/Description	MS	MSD	Matrix	Date Collected	Time Collected	VOA	BNA	Pest/PCB	Herb	Metal	CN	Sulfide	IC Anions	NO2 NO3
001	BOWCP8			W	9-15-99	0650	✓	✓			✓		✓	✓	✓
002	BOWCP9			↓	↓	0518	✓								

Special Instructions:
saf # B99-085

COMPOSITE WASTE

DATE/REVISIONS:
1. OGCSC = 1-propanol, Ethanol
2. Met① = As, Ba, Cd, Cr, Pb, Se, Ag, Cu,
3. Ni, V, Zn, Be
4. Ang① = ICCL, ICFL, ICNO2, ICNO3, ICPO4,
5. ICSO4, 1PH
6. Run matrix QC

Relinquished by	Received by	Date	Time
Fed Ex	TMurray	9-17-99	1020

Relinquished by	Received by	Date	Time
	ORIGINAL REWRITTEN		

RECRA LabNet Use Only

Samples were:
1) Shipped ✓ or Hand Delivered ___
Airbill # *
2) Ambient or (Chilled)
3) Received in Good Condition (Y) or N
4) Labels Indicate Properly Preserved (Y) or N
5) Received Within Holding Times Y or (N) — IC, pH autofhold

COC Tape was:
1) Present on Outer Package (Y) or N
2) Unbroken on Outer Package (Y) or N
3) Present on Sample (Y) or N
4) Unbroken on Sample (Y) or N
COC Record Present Upon Sample Rec't (Y) or N
Cooler Temp. 4.1 °C

Discrepancies Between Samples Labels and COC Record? (Y) or N
NOTES: Sulfide bottle = 1L not 500mL as chain indicates

*4235795295(?)

Bechtel Hanford Inc.		CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST	B99-085-04	Page 1 of 1

Collector	Company Contact / Telephone No.	Project Coordinator	Price Code	Data Turnaround
Bowers/Trice	C Cearlock 372-9574	TRENT, SJ	7N	45 Days
Project Designation	**Sampling Location**	**SAF No.**		
200 Area Source characterization - 200-CW-1 OU - QC Sa	200 East	B99-085		
Ice Chest No.	**Field Logbook No.**	**Method of Shipment**		
ERC 96 024	EL1511	Fed Ex		
Shipped To	**Offsite Property No.**	**Bill of Lading/Air Bill No.**		
TMA/RECRA DTD 9-15-99	A990259	4285795295 61		
		COA B20CW167C		

POSSIBLE SAMPLE HAZARDS/REMARKS	Preservation	ZnAc+NaOH to pH >9 Cool	Cool 4C	H2SO4 to pH <2 Cool 4C	Cool 4C	HNO3 to pH <2	HCl to pH <2 Cool 4C	HNO3 to pH <2			
	Type of Container	P	P	P	aG	P	aGs*	P			
	No. of Container(s)	1	1	1	2	2	3	3			
Special Handling and/or Storage	Volume	500mL	1000mL	1000mL	1000mL	1000mL	40mL	500mL			
SAMPLE ANALYSIS		Sulfides - 9030	See item (1) in Special Instructions	NO2/NO3 - 353.1; Ammonia - 350.3	Semi-VOA - 8270A (TCL)	Gross Alpha, Gross Beta	VOA - 8260A (TCL); VOA - 8260A (Add-On) {1-Propanol, Ethanol}	See item (2) in Special Instructions			

Sample No.	Matrix *	Sample Date	Sample Time										
B0WCP8	Water	9.15.99	0650	X	X	X	X		X	X			
B0WCP9	Water	9.15.99	0518						X				

CHAIN OF POSSESSION	Sign/Print Names	Date/Time
Relinquished By	Doug Bowers	9-17-99/1530
Received By	Ref 1B	9-17-99/1730
Relinquished By	Ref 1B	91699 1300
Received By	SJ Gale	91699 1300
Relinquished By	SJ Gale	91699 1300
Received By	FED EX	
Relinquished By	Fed Ex	9-17-99 1020
Received By	TM Murray	9-17-99 1020

SPECIAL INSTRUCTIONS

(1) IC Anions - 300.0 {Chloride, Fluoride, Nitrate, Nitrite, Phosphate, Sulfate}; pH (Water) - 9040
(2) ICP Metals - 6010A (Supertrace) {Arsenic, Barium, Cadmium, Chromium, Lead, Selenium, Silver}; ICP Metals - 6010A (Supertrace Add-On) {Copper, Nickel, Vanadium, Zinc} Be
samples from non rad area
COLLECTOR UNAVAILABLE TO SIGN COC.

Matrix *
Soil
Water
Vapor
Other Solid
Other Liquid

LABORATORY SECTION	Received By	Title	Date/Time
FINAL SAMPLE DISPOSITION	Disposal Method	Disposed By	Date/Time

RECRA LabNet
a division of Recra Environmental, Inc.
Virtual Laboratories Everywhere



Recra LabNet Philadelphia
Analytical Report

Client : TNU-HANFORD B99-085
RFW# : 9909L126
SDG/SAF# : H0535/B99-085

W.O.# : 10985-001-001-9999-00
Date Received: 09-17-99

METALS CASE NARRATIVE

1. This narrative covers the analyses of 1 water sample.

2. The sample was prepared and analyzed in accordance with methods checked on the attached glossary.

3. All analyses were performed within the required holding times.

4. The cooler temperature has been recorded on the Chain of Custody.

5. All Initial and Continuing Calibration Verifications (ICV/CCVs) were within the 90-110% control limits (80-120% for Mercury).

6. All Initial and Continuing Calibration Blanks (ICB/CCBs) were within control limits (less than the PQL).

7. All preparation/method blanks (MB) were within method criteria {less than the Practical Quantitation Limit (3X the IDL) or samples greater than 20X MB value}. Refer to the Inorganics Method Blank Data Summary.

8. All ICP Interference Check Standards were within control limits.

9. All laboratory control samples (LCS) were within the laboratory control limits. Refer to the Inorganics Laboratory Control Standards Report.

10. All matrix spike (MS) recoveries were within the 75-125% control limits. Refer to the Inorganics Accuracy Report.

11. The duplicate analyses for 3 analytes were outside the 20% Relative Percent Difference (RPD) control limits. Refer to the Inorganics Precision Report.

12. For the purposes of this report, the data has been reported to the Instrument Detection Limit (IDL). Values between the IDL and the Practical Quantitation Limit (PQL) are acquired in a region of less-certain quantification.

J. Michael Taylor
Vice President
Philadelphia Analytical Laboratory

10-19-99
Date

mld/m09-126

METALS METHOD GLOSSARY

The following methods are used as reference for the digestion and analysis of samples contained within this Recra Lot#: 9909L126

Leaching Procedure: __1310 __1311 __1312 __Other:_____

CLP Metals __ Digestion and __ Analysis Methods: __ILM03.0 __ILM04.0

Metals Digestion Methods: ✓3005A __3010A __3015 __3020A __3050A __3051 __200.7 __SS17
__Other: _____

Metals Analysis Methods

	SW846	EPA	STD MTD	EPA OSWR	USATHAMA
Aluminum	__6010B	__200.7			__99
Antimony	__6010B __7041 [5]	__200.7 __204.2			__99
Arsenic	✓6010B __7060A [5]	__200.7 __206.2	__3113B		__99
Barium	✓6010B	__200.7			__99
Beryllium	✓6010B	__200.7			__99
Bismuth	__6010B [1]	__200.7 [1]		__1620	__99
Boron	__6010B	__200.7			__99
Cadmium	✓6010B __7131A [5]	__200.7 __213.2			__99
Calcium	__6010B	__200.7			__99
Chromium	✓6010B __7191 [5]	__200.7 __218.2			__SS17
Cobalt	__6010B	__200.7			__99
Copper	✓6010B __7211 [5]	__200.7 __220.2			__99
Iron	__6010B	__200.7			__99
Lead	✓6010B __7421 [5]	__200.7 __239.2	__3113B		__99
Lithium	__6010B __7430 [4]	__200.7		__1620	__99
Magnesium	__6010B	__200.7			__99
Manganese	__6010B	__200.7			__99
Mercury	__7470A [3] __7471A [3]	__245.1 [2] __245.5 [2]			__99
Molybdenum	__6010B	__200.7			__99
Nickel	✓6010B	__200.7			__99
Potassium	__6010B __7610 [4]	__200.7 __258.1 [4]			__99
Rare Earths	__6010B [1]	__200.7 [1]		__1620	__99
Selenium	✓6010B __7740 [5]	__200.7 __270.2	__3113B		__99
Silicon	__6010B [1]	__200.7		__1620	__99
Silica	__6010B	__200.7		__1620	__99
Silver	✓6010B __7761 [5]	__200.7 __272.2			__99
Sodium	__6010B __7770 [4]	__200.7 __273.1 [4]			__99
Strontium	__6010B	__200.7			__99
Thallium	__6010B __7841 [5]	__200.7 __279.2 __200.9			__99
Tin	__6010B	__200.7			__99
Titanium	__6010B	__200.7			__99
Uranium	__6010B [1]	__200.7 [1]		__1620	__99
Vanadium	✓6010B	__200.7			__99
Zinc	✓6010B	__200.7			__99
Zirconium	__6010B [1]	__200.7 [1]		__1620	__99

Other:_____ Method:_____

METHOD REFERENCES AND DATA QUALIFIERS

DATA QUALIFIERS

U = Indicates that the parameter was not detected at or above the reported limit. The associated numerical value is the sample detection limit.

* = Indicates that the original sample result is greater than 4x the spike amount added.

ABBREVIATIONS

MB = Method or Preparation Blank.
MS = Matrix Spike.
MSD = Matrix Spike Duplicate.
REP = Sample Replicate
LCS = Laboratory Control Sample.
NC = Not calculated.

ANALYTICAL METAL METHODS

1. Not included in the method element list.

2. Modified Hg: Hg1 and Hg2 require less total volume of digestate due to the autosampler analysis. Sample volumes and reagents for mercury determinations in water and soil have been proportionately scaled down to adapt to this semi-automated technique. The sample volume used for water analysis is 33 mL. For soils, 0.1 grams of sample is taken to a final volume of 50 mL (including all reagents).

3. Modified Hg: Hg1 and Hg2 require less total volume of digestate due to the autosampler analysis. Sample volumes and reagents for mercury determinations in water and soil have been proportionately scaled down to adapt to this semi-automated technique. The sample volume used for water analysis is 33 mL. For soils, three 0.1 gram of sample is taken to a final volume of 50 mL (including all reagents).

4. Flame AA.

5. Graphite Furnace AA.

Recra LabNet - Lionville

INORGANICS DATA SUMMARY REPORT 10/19/99

CLIENT: TNU-HANFORD B99-085 RECRA LOT #: 9909L126

WORK ORDER: 10985-001-001-9999-00

SAMPLE	SITE ID	ANALYTE	RESULT	UNITS	REPORTING LIMIT	DILUTION FACTOR
-001	B0WCP8	Silver, Total	1.0 u	UG/L	1.0	1.0
		Arsenic, Total	3.3 u	UG/L	3.3	1.0
		Barium, Total	0.39	UG/L	0.30	1.0
		Beryllium, Total	0.10 u	UG/L	0.10	1.0
		Cadmium, Total	0.30 u	UG/L	0.30	1.0
		Chromium, Total	0.80 u	UG/L	0.80	1.0
		Copper, Total	1.2 u	UG/L	1.2	1.0
		Nickel, Total	1.2 u	UG/L	1.2	1.0
		Lead, Total	2.1 u	UG/L	2.1	1.0
		Selenium, Total	3.7 u	UG/L	3.7	1.0
		Vanadium, Total	0.60 u	UG/L	0.60	1.0
		Zinc, Total	0.99	UG/L	0.80	1.0

Recra LabNet - Lionville

INORGANICS METHOD BLANK DATA SUMMARY PAGE 10/19/99

CLIENT: TNU-HANFORD B99-085 RECRA LOT #: 9909L126

WORK ORDER: 10985-001-001-9999-00

SAMPLE	SITE ID	ANALYTE	RESULT	UNITS	REPORTING LIMIT	DILUTION FACTOR
BLANK1	99L0682-MB1	Silver, Total	1.0 u	UG/L	1.0	1.0
		Arsenic, Total	3.3 u	UG/L	3.3	1.0
		Barium, Total	0.37	UG/L	0.30	1.0
		Beryllium, Total	0.10 u	UG/L	0.10	1.0
		Cadmium, Total	0.30 u	UG/L	0.30	1.0
		Chromium, Total	0.80 u	UG/L	0.80	1.0
		Copper, Total	1.2 u	UG/L	1.2	1.0
		Nickel, Total	1.2 u	UG/L	1.2	1.0
		Lead, Total	2.1 u	UG/L	2.1	1.0
		Selenium, Total	3.7 u	UG/L	3.7	1.0
		Vanadium, Total	0.60 u	UG/L	0.60	1.0
		Zinc, Total	1.5	UG/L	0.80	1.0

Recra LabNet - Lionville

INORGANICS ACCURACY REPORT 10/19/99

CLIENT: TNU-HANFORD B99-085 RECRA LOT #: 9909L126

WORK ORDER: 10985-001-001-9999-00

SAMPLE	SITE ID	ANALYTE	SPIKED SAMPLE	INITIAL RESULT	SPIKED AMOUNT	%RECOV	DILUTION FACTOR(SPK)
-001	B0WCP8	Silver, Total	49.8	1.0 u	50.0	99.6	1.0
		Arsenic, Total	2050	3.3 u	2000	102.5	1.0
		Barium, Total	1980	0.39	2000	99.0	1.0
		Beryllium, Total	50.2	0.10u	50.0	100.4	1.0
		Cadmium, Total	51.0	0.30u	50.0	102.0	1.0
		Chromium, Total	202	0.80u	200	101.2	1.0
		Copper, Total	251	1.2 u	250	100.2	1.0
		Nickel, Total	498	1.2 u	500	99.6	1.0
		Lead, Total	510	2.1 u	500	102.1	1.0
		Selenium, Total	2030	3.7 u	2000	101.6	1.0
		Vanadium, Total	511	0.60u	500	102.3	1.0
		Zinc, Total	501	0.99	500	100.1	1.0

Recra LabNet - Lionville

INORGANICS PRECISION REPORT 10/19/99

CLIENT: TNU-HANFORD B99-085 RECRA LOT #: 9909L126

WORK ORDER: 10985-001-001-9999-00

SAMPLE	SITE ID	ANALYTE	INITIAL RESULT	REPLICATE	RPD	DILUTION FACTOR(REP)
-001REP	B0WCP8	Silver, Total	1.0 u	1.0 u	NC	1.0
		Arsenic, Total	3.3 u	3.3 u	NC	1.0
		Barium, Total	0.39	1.1	95.3	1.0
		Beryllium, Total	0.10u	0.10u	NC	1.0
		Cadmium, Total	0.30u	0.30u	NC	1.0
		Chromium, Total	0.80u	0.80u	NC	1.0
		Copper, Total	1.2 u	1.2 u	NC	1.0
		Nickel, Total	1.2 u	1.2 u	NC	1.0
		Lead, Total	2.1 u	2.1 u	NC	1.0
		Selenium, Total	3.7 u	3.7 u	NC	1.0
		Vanadium, Total	0.60u	0.65	~~NC~~ 200	1.0
		Zinc, Total	0.99	1.7	52.8	1.0

Correction AD 10/19/99

Recra LabNet - Lionville

INORGANICS LABORATORY CONTROL STANDARDS REPORT 10/19/99

CLIENT: TNU-HANFORD B99-085 RECRA LOT #: 9909L126

WORK ORDER: 10985-001-001-9999-00

SAMPLE	SITE ID	ANALYTE	SPIKED SAMPLE	SPIKED AMOUNT	UNITS	%RECOV
LCS1	99L0682-LC1	Silver, LCS	497	500	UG/L	99.4
		Arsenic, LCS	10200	10000	UG/L	101.9
		Barium, LCS	5020	5000	UG/L	100.4
		Beryllium, LCS	251	250	UG/L	100.4
		Cadmium, LCS	249	250	UG/L	99.8
		Chromium, LCS	499	500	UG/L	99.8
		Copper, LCS	1250	1250	UG/L	99.7
		Nickel, LCS	2000	2000	UG/L	100
		Lead, LCS	2470	2500	UG/L	98.9
		Selenium, LCS	10200	10000	UG/L	101.8
		Vanadium, LCS	2520	2500	UG/L	100.9
		Zinc, LCS	998	1000	UG/L	99.8

Recra LabNet - Lionville Laboratory
INORGANIC ANALYTICAL DATA PACKAGE FOR
TNU-HANFORD B99-085

DATE RECEIVED: 09/17/99 RFW LOT # :9909L126

CLIENT ID /ANALYSIS	RFW #	MTX	PREP #	COLLECTION	EXTR/PREP	ANALYSIS
B0WCP8						
SILVER, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
SILVER, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
SILVER, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
ARSENIC, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
ARSENIC, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
ARSENIC, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
BARIUM, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
BARIUM, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
BARIUM, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
BERYLLIUM, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
BERYLLIUM, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
BERYLLIUM, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
CADMIUM, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
CADMIUM, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
CADMIUM, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
CHROMIUM, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
CHROMIUM, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
CHROMIUM, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
COPPER, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
COPPER, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
COPPER, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
NICKEL, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
NICKEL, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
NICKEL, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
LEAD, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
LEAD, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
LEAD, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
SELENIUM, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
SELENIUM, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
SELENIUM, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
VANADIUM, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
VANADIUM, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99
VANADIUM, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99
ZINC, TOTAL	001	W	99L0682	09/15/99	10/06/99	10/15/99
ZINC, TOTAL	001 REP	W	99L0682	09/15/99	10/06/99	10/15/99

Recra LabNet - Lionville Laboratory
INORGANIC ANALYTICAL DATA PACKAGE FOR
TNU-HANFORD B99-085

DATE RECEIVED: 09/17/99 RFW LOT # :9909L126

CLIENT ID /ANALYSIS	RFW #	MTX	PREP #	COLLECTION	EXTR/PREP	ANALYSIS
ZINC, TOTAL	001 MS	W	99L0682	09/15/99	10/06/99	10/15/99

LAB QC:

SILVER LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
SILVER, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
ARSENIC LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
ARSENIC, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
BARIUM LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
BARIUM, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
BERYLLIUM LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
BERYLLIUM, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
CADMIUM LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
CADMIUM, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
CHROMIUM LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
CHROMIUM, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
COPPER LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
COPPER, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
NICKEL LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
NICKEL, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
LEAD LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
LEAD, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
SELENIUM LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
SELENIUM, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
VANADIUM LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
VANADIUM, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99
ZINC LABORATORY	LC1 BS	W	99L0682	N/A	10/06/99	10/08/99
ZINC, TOTAL	MB1	W	99L0682	N/A	10/06/99	10/08/99

RECRA LabNet Use Only
9909L126

Custody Transfer Record/Lab Work Request

Page 1 of 1

RECRA LabNet

A11 FIELD PERSONNEL: COMPLETE ONLY SHADED AREAS

Client: TDU Hanford B99-085
Est. Final Proj. Sampling Date:
Project #: 10985-001-001-9999-00
Project Contact/Phone #:
RECRA Project Manager: OJ
QC: spec Del: Std TAT: 30 day
Date Rec'd: 9-17-99 Date Due: 10/17/99
Account #:

	VOA	BNA	Pest/PCB	Herb					Metal	CN	Sulfide	IC Anions	Free NO3			
Refrigerator #	1	6							6		6	6	6			
#/Type Container Liquid	3V	2AG							3P		1P	1P	1P			
#/Type Container Solid																
Volume Liquid	40ml	1L							1L		1L	1L	1L			
Volume Solid											ZNAC					
Preservatives									HNO3		NaOH	–	H2SO4			
RECRA LabNet Use Only	0024H OGCSC	0025H 8020H	8270						MET①		ISFD	ANG①	IN3N2			

MATRIX CODES: S - Soil; SE - Sediment; SO - Solid; SL - Sludge; W - Water; O - Oil; A - Air; DS - Drum Solids; DL - Drum Liquids; L - EP/TCLP Leachate; WI - Wipe; X - Other; F - Fish

Lab ID	Client ID/Description	MS	MSD	Matrix	Date Collected	Time Collected	VOA	BNA	Pest/PCB	Herb					Metal	CN	Sulfide	IC Anions	Free NO3
001	B0WCP8			W	9-15-99	0650	✓	✓							✓		✓	✓	✓
002	B0WCP9			↓	↓	0518	✓												

Special Instructions:
saf # B99-085

COMPOSITE WASTE

DATE/REVISIONS:
1. OGCSC = 1-propanol, Ethanol
2. MET① = As, Ba, Cd, Cr, Pb, Se, Ag, Cu,
3. Ni, V, Zn, Be
4. ANG① = ICCL, ICFL, ICNO2, ICNO3, ICPO4,
5. ICSO4, IPH
6. Run matrix QC

RECRA LabNet Use Only

Samples were: 1) Shipped ✓ or Hand Delivered ___
Airbill # *
2) Ambient or (Chilled)
3) Received in Good Condition (Y) or N
4) Labels Indicate Properly Preserved (Y) or N
5) Received Within Holding Times IC, pH a.f.f.h.o.d. Y or (N)

COC Tape was:
1) Present on Outer Package (Y) or N
2) Unbroken on Outer Package (Y) or N
3) Present on Sample (Y) or N
4) Unbroken on Sample (Y) or N
COC Record Present Upon Sample Rec'd (Y) or N
Cooler Temp 4.1 °C

Relinquished by	Received by	Date	Time
Fed Ex	T Murray	9-17-99	1020

Relinquished by	Received by	Date	Time
	ORIGINAL REWRITTEN		

Discrepancies Between Samples Labels and COC Record? (Y) or N
NOTES: Sulfide bottle = 1L not 500 mL as chain indicates

Bechtel Hanford Inc.		CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST	B99-085-04	Page 1 of 1

Collector	Company Contact / Telephone No.	Project Coordinator	Price Code	Data Turnaround
Bowers/Trice	C Cearlock 372-9574	TRENT, SJ	7N	45 Days
Project Designation	**Sampling Location**	**SAF No.**		
200 Area Source characterization - 200-CW-1 OU - QC Sa	200 East	B99-085		
Ice Chest No.	**Field Logbook No.**	**Method of Shipment**		
ERC 96 024	FLT11	Fed Ex		
Shipped To	**Offsite Property No.**	**Bill of Lading/Air Bill No.**		
TMA/RECRA DAD 9-15-99	A990259	4285 7952 9561		
		COA B20CW106 7C		

POSSIBLE SAMPLE HAZARDS/REMARKS	Preservation	ZnAc+NaOH to pH >9 Cool	Cool 4C	H2SO4 to pH <2 Cool 4C	Cool 4C	HNO3 to pH <2	HCl to pH <2 Cool 4C	HNO3 to pH <2			
	Type of Container	P	P	P	aG	P	aGs*	P			
	No. of Container(s)	1	1	1	2	2	3	3			
Special Handling and/or Storage	Volume	500mL	1000mL	1000mL	1000mL	1000mL	40mL	500mL			
SAMPLE ANALYSIS		Sulfides - 9030	See item (1) in Special Instructions.	NO2/NO3 - 353.1; Ammonia - 350.3	Semi-VOA - 8270A (TCL)	Gross Alpha; Gross Beta	VOA - 8260A (TCL); VOA - 8260A (Add-On) {1-Propanol, Ethanol}	See item (2) in Special Instructions.			

Sample No.	Matrix *	Sample Date	Sample Time									
B0WCP8	Water	9.15.99	0650	X	X	X	X		X	X		
B0WCP9	Water	9.15.99	0518						X			

CHAIN OF POSSESSION	Sign/Print Names	Date/Time
Relinquished By	Doug Bowers	9-17-99/1530
Received By	Ref 1B	9-17-99/1530
Relinquished By	Ref 1B	91699 1300
Received By	SJ Gale SJ Gale	91699 1300
Relinquished By	SJ Gale SJ Gale	91699 1300
Received By	FED EX	
Relinquished By	FED EX	9-17-99 1020
Received By	TM Murray	9-17-99 1020

SPECIAL INSTRUCTIONS

(1) IC Anions - 300.0 {Chloride, Fluoride, Nitrate, Nitrite, Phosphate, Sulfate}; pH (Water) - 9040
(2) ICP Metals - 6010A (Supertrace) {Arsenic, Barium, Cadmium, Chromium, Lead, Selenium, Silver}; ICP Metals - 6010A (Supertrace Add-On) {Copper, Nickel, Vanadium, Zinc} Be

samples from near prod area

COLLECTOR UNAVAILABLE TO SIGN COC.

Matrix *
Soil
Water
Vapor
Other Solid
Other Liquid

LABORATORY SECTION	Received By	Title	Date/Time
FINAL SAMPLE DISPOSITION	Disposal Method	Disposed By	Date/Time

RECRA LabNet
a division of Recra Environmental, Inc.
Virtual Laboratories Everywhere

Recra LabNet Philadelphia
Analytical Report

Client : TNU-HANFORD B99-085
RFW# : 9909L126
SDG# : H0535
SAF# : B99-085

W.O. # : 10985-001-001-9999-00
Date Received: 09-17-99

INORGANIC CASE NARRATIVE

1. This narrative covers the analyses of 1 water sample.

2. The sample was prepared and analyzed in accordance with the methods checked on the attached glossary.

3. Sample holding times as required by the method and/or contract were met with the exception of pH, Nitrate, Nitrite and Phosphate which were received past hold.

4. The cooler temperature was recorded on the chain-of-custody.

5. The method blanks were within method criteria.

6. The Laboratory Control Samples (LCS) were within the laboratory control limits. The duplicate LCS were within the 20% Relative Percent Difference (RPD) control limit.

7. The matrix spike recoveries were within the 75-125% control limits. The matrix spike duplicates were within the 20% RPD control limit.

8. The replicate analyses were within the 20% RPD control limit.

Pat Ec

J. Michael Taylor
Vice President
Philadelphia Analytical Laboratory

10-18-99
Date

njp\i09-126

Recra LabNet Philadelphia

WET CHEMISTRY

METHODS GLOSSARY FOR WATER SAMPLE ANALYSIS

	EPA /600	SW846	OTHER
Acidity	___ 305.1		
___Alkalinity ___Bicarbonate ___Carbonate	___ 310.1		
BOD	___ 405.1		___ 5210B (b)
Ion Chormatography:			
___Bromide ✓ Chloride ✓ Fluoride	✓ 300.0	___ 9056	
✓ Nitrite ___Nitrate ✓ Phosphate	✓ 300.0	___ 9056	
✓ Sulfate ___Formate ___Acetate ___Oxalate	✓ 300.0	___ 9056	
Chloride	___ 325.2	___ 9251	
Chorine, Residual	___ 330.5 (mod)		
Cyanide, Amenable to Chorination	___ 335.2	___ 9010B	
Cyanide, Total	___ 335.2	___ 9010B ___ 9014	___ ILMO4.0 (e)
Cyanide, Weak Acid Dissociable		___	___ 412 (a) ___ 4500CN-I (b
COD	___ 410.4(mod)		___ 5220C (b)
Color	___ 110.2		
Corrosivity by Coupon		___ 1110(mod)	
Chromium VI		___ 7196A	___ 3500Cr-D (b)
Fluoride	___ 340.2		___ 4500-FC
Hardness, Calcium	___ 215.2		
Hardness, Total	___ 130.2		
Iodide			___ ASTM D19P202 (1)
Surfactant	___ 425.1		
✓ Nitrate-Nitrite ___Nitrate ___Nitrite	✓ 353.2		
Ammonia	✓ 350.3		
Total___ Kjeldahl ___ Organic Nitrogen	___ 351.4		
Total___ Organic ___ Inorganic Carbon	___ 415.1	___ 9060	
Oil & Grease	___ 413.1	___ 9070	
✓ pH ____ pH; paper	___ 150.1	✓ 9040B ___ 9041A	
Petroleum Hydrocarbons, Total Recoverable	___ 418.1		
Phenol	___ 420.1 ___ 420.2	___ 9065 ___ 9066	
___Ortho ___Total Phosphate	___ 365.2		___ 4500-P B ___ C
Salinity			___ 210A (a) ___ 2520 (b)
Settleable Solids	___ 160.5		
Sulfide	___ 376.1 ___ 376.2	✓ 9030B/9034 (acid soluble)	
Reactive ___Cyanide ___Sulfide		___ Section 7.3	
Silica	___ 370.1		
Sulfite	___ 377.1		
Sulfate	___ 375.4	___ 9038	
Specific Conductance	___ 120.1	___ 9050A	
Specific Gravity			___ D5057-90 ___ 213E (a)
Synthetic Precipitation Leach		___ 1312	
Total ___Dissolved ___Suspended ___Solids	160 ___.1 ___.2 ___.3		
Total Organic Halides	___ 450.1	___ 9020B	
Turbidity	___ 180.1		
Volatile Solids:			
___Total ___Dissolved ___Suspended	___ 160.4		

Other: **Method:**

Recra LabNet Philadelphia
METHOD REFERENCES AND DATA QUALIFIERS

DATA QUALIFIERS

U = Indicates that the parameter was not detected at or above the reported limit. The associated numerical value is the sample detection limit.

* = Indicates that the original sample result is greater than 4x the spike amount added.

ABBREVIATIONS

MB = Method or Preparation Blank.
MS = Matrix Spike.
MSD = Matrix Spike Duplicate.
REP = Sample Replicate
LC = Laboratory Control Sample.
NC = Not calculated.

A suffix of -R, -S, or -T following these codes indicate a replicate, spike or sample duplicate analysis respectively.

ANALYTICAL WET CHEMISTRY METHODS

1. ASTM Standard Methods.

2. USEPA Methods for Chemical Analysis of Water and Wastes (USEPA 600/4-79-020).

3. Test Methods for Evaluating Solid Waste (USEPA SW-846).

a. Standard Methods for the Examination of Water and Waste, 16 ed, (1983).

b. Standard Methods for the Examination of Water and Waste, 17 ed, (1989)/18ed (1992).

c. Method of Soil Analysis, Part 1, Physical and Mineralogical Methods, 2nd ed, (1986).

d. Method of Soil Analysis, Part 2, Chemical and Microbiological Properties, Am. Soc. Agron., Madison, WI (1965).

e. USEPA Contract Laboratory Program, Statement of Work for Inorganic Analysis.

f. Code of Federal Regulations.

Recra LabNet - Lionville

INORGANICS DATA SUMMARY REPORT 10/04/99

CLIENT: TNU-HANFORD B99-085 RECRA LOT #: 9909L126

WORK ORDER: 10985-001-001-9999-00

SAMPLE	SITE ID	ANALYTE	RESULT	UNITS	REPORTING LIMIT	DILUTION FACTOR
-001	B0WCP8	Chloride by IC	0.25 u	MG/L	0.25	1.0
		Fluoride by IC	0.50 u	MG/L	0.50	1.0
		Nitrite by IC	0.25 u	MG/L	0.25	1.0
		Nitrate by IC	0.25 u	MG/L	0.25	1.0
		Phosphate by IC	0.25 u	MG/L	0.25	1.0
		Sulfate by IC	0.25 u	MG/L	0.25	1.0
		Nitrate Nitrite	0.02 u	MG-N/L	0.02	1.0
		Ammonia, as N	0.10 u	MG/L	0.10	1.0
		pH	6.5	PH UNITS	0.01	1.0
		Sulfide	1.0 u	MG/L	1.0	1.0

Recra LabNet - Lionville

INORGANICS METHOD BLANK DATA SUMMARY PAGE 10/04/99

CLIENT: TNU-HANFORD B99-085 RECRA LOT #: 9909L126

WORK ORDER: 10985-001-001-9999-00

SAMPLE	SITE ID	ANALYTE	RESULT	UNITS	REPORTING LIMIT	DILUTION FACTOR
BLANK10	99LICB79-MB1	Chloride by IC	0.25 u	MG/L	0.25	1.0
		Fluoride by IC	0.50 u	MG/L	0.50	1.0
		Nitrite by IC	0.25 u	MG/L	0.25	1.0
		Nitrate by IC	0.25 u	MG/L	0.25	1.0
		Sulfate by IC	0.25 u	MG/L	0.25	1.0
BLANK10	99LICC79-MB1	Phosphate by IC	0.25 u	MG/L	0.25	1.0
BLANK10	99LN3A47-MB1	Nitrate Nitrite	0.02 u	MG-N/L	0.02	1.0
BLANK10	99LAMA36-MB1	Ammonia, as N	0.10 u	MG/L	0.10	1.0
BLANK10	99LSD047-MB1	Sulfide	1.0 u	MG/L	1.0	1.0

Recra LabNet - Lionville

INORGANICS ACCURACY REPORT 10/04/99

CLIENT: TNU-HANFORD B99-085 RECRA LOT #: 9909L126

WORK ORDER: 10985-001-001-9999-00

SAMPLE	SITE ID	ANALYTE	SPIKED SAMPLE	INITIAL RESULT	SPIKED AMOUNT	%RECOV	DILUTION FACTOR(SPK)
-001	B0WCP8	Chloride by IC	5.1	0.00	5.0	101.9	1.0
		Fluoride by IC	10.9	0.00	10.0	109.3	1.0
		Nitrite by IC	5.3	0.25u	5.0	106.1	1.0
		Nitrate by IC	5.1	0.25u	5.0	101.7	1.0
		Phosphate by IC	5.2	0.25u	5.0	103.1	1.0
		Sulfate by IC	5.1	0.25u	5.0	102.2	1.0
		Nitrate Nitrite	0.50	0.02u	0.50	100.8	1.0
		Nitrate Nitrite MSD	0.50	0.02u	0.50	100.2	1.0
		Ammonia, as N	0.95	0.10u	1.0	95.3	1.0
		Sulfide	9.7	0.00	9.9	98.0	1.0
		Sulfide MSD	9.6	0.00	9.9	97.0	1.0
BLANK10	99LICB79-MB1	Chloride by IC	4.9	0.25u	5.0	97.6	1.0
		Fluoride by IC	10.6	0.50u	10.0	105.7	1.0
		Nitrite by IC	4.9	0.25u	5.0	98.4	1.0
		Nitrate by IC	4.9	0.25u	5.0	98.0	1.0
		Sulfate by IC	4.8	0.25u	5.0	96.3	1.0
BLANK10	99LICC79-MB1	Phosphate by IC	5.0	0.25u	5.0	99.1	1.0
BLANK10	99LN3A47-MB1	Nitrate Nitrite	0.50	0.02u	0.50	101.0	1.0
		Nitrate Nitrite MSD	0.51	0.02u	0.50	102.6	1.0
BLANK10	99LAMA36-MB1	Ammonia, as N	1.0	0.10u	1.0	104.0	1.0
		Ammonia, as N MSD	1.0	0.10u	1.0	103.0	1.0
BLANK10	99LSD047-MB1	Sulfide	9.9	1.0 u	9.9	100	1.0
		Sulfide MSD	10.0	1.0 u	9.9	101.0	1.0

Recra LabNet - Lionville

INORGANICS DUPLICATE SPIKE REPORT 10/04/99

CLIENT: TNU-HANPORD B99-085 RECRA LOT #: 9909L126

WORK ORDER: 10985-001-001-9999-00

SAMPLE	SITE ID	ANALYTE	SPIKE#1 %RECOV	SPIKE#2 %RECOV	%DIFF
-001	B0WCP8	Nitrate Nitrite	100.8	100.2	0.60
		Sulfide	98.0	97.0	1.0
BLANK10	99LN3A47-MB1	Nitrate Nitrite	101.0	102.6	1.6
BLANK10	99LAMA36-MB1	Ammonia, as N	104.0	103.0	0.97
BLANK10	99LSD047-MB1	Sulfide	100	101.0	1.0

Recra LabNet - Lionville

INORGANICS PRECISION REPORT 10/04/99

CLIENT: TNU-HANFORD B99-085 RECRA LOT #: 9909L126

WORK ORDER: 10985-001-001-9999-00

SAMPLE	SITE ID	ANALYTE	INITIAL RESULT	REPLICATE	RPD	DILUTION FACTOR(REP)
-001REP	B0WCP8	Chloride by IC	0.25u	0.25u	NC	1.0
		Fluoride by IC	0.50u	0.50u	NC	1.0
		Nitrite by IC	0.25u	0.25u	NC	1.0
		Nitrate by IC	0.25u	0.25u	NC	1.0
		Phosphate by IC	0.25u	0.25u	NC	1.0
		Sulfate by IC	0.25u	0.25u	NC	1.0
		Nitrate Nitrite	0.02u	0.02u	NC	1.0
		Ammonia, as N	0.10u	0.10u	NC	1.0
		Sulfide	1.0 u	1.0 u	NC	1.0

Recra LabNet - Lionville Laboratory
INORGANIC ANALYTICAL DATA PACKAGE FOR
TNU-HANFORD B99-085

DATE RECEIVED: 09/17/99 RFW LOT # :9909L126

CLIENT ID /ANALYSIS	RFW #	MTX	PREP #	COLLECTION	EXTR/PREP	ANALYSIS
B0WCP8						
CHLORIDE BY IC	001	W	99LICB79	09/15/99	09/23/99	09/23/99
CHLORIDE BY IC	001 REP	W	99LICB79	09/15/99	09/23/99	09/23/99
CHLORIDE BY IC	001 MS	W	99LICB79	09/15/99	09/23/99	09/23/99
FLUORIDE BY IC	001	W	99LICB79	09/15/99	09/23/99	09/23/99
FLUORIDE BY IC	001 REP	W	99LICB79	09/15/99	09/23/99	09/23/99
FLUORIDE BY IC	001 MS	W	99LICB79	09/15/99	09/23/99	09/23/99
NITRITE BY IC	001	W	99LICB79	09/15/99	09/23/99	09/23/99
NITRITE BY IC	001 REP	W	99LICB79	09/15/99	09/23/99	09/23/99
NITRITE BY IC	001 MS	W	99LICB79	09/15/99	09/23/99	09/23/99
NITRATE BY IC	001	W	99LICB79	09/15/99	09/23/99	09/23/99
NITRATE BY IC	001 REP	W	99LICB79	09/15/99	09/23/99	09/23/99
NITRATE BY IC	001 MS	W	99LICB79	09/15/99	09/23/99	09/23/99
PHOSPHATE BY IC	001	W	99LICC79	09/15/99	09/23/99	09/23/99
PHOSPHATE BY IC	001 REP	W	99LICC79	09/15/99	09/23/99	09/23/99
PHOSPHATE BY IC	001 MS	W	99LICC79	09/15/99	09/23/99	09/23/99
SULFATE BY IC	001	W	99LICB79	09/15/99	09/23/99	09/23/99
SULFATE BY IC	001 REP	W	99LICB79	09/15/99	09/23/99	09/23/99
SULFATE BY IC	001 MS	W	99LICB79	09/15/99	09/23/99	09/23/99
NITRATE NITRITE	001	W	99LN3A47	09/15/99	10/01/99	10/01/99
NITRATE NITRITE	001 REP	W	99LN3A47	09/15/99	10/01/99	10/01/99
NITRATE NITRITE	001 MS	W	99LN3A47	09/15/99	10/01/99	10/01/99
NITRATE NITRITE	001 MSD	W	99LN3A47	09/15/99	10/01/99	10/01/99
AMMONIA	001	W	99LAMA36	09/15/99	09/24/99	09/24/99
AMMONIA	001 REP	W	99LAMA36	09/15/99	09/24/99	09/24/99
AMMONIA	001 MS	W	99LAMA36	09/15/99	09/24/99	09/24/99
PH	001	W	99LPH102	09/15/99	09/23/99	09/23/99
SULFIDE	001	W	99LSD047	09/15/99	09/19/99	09/20/99
SULFIDE	001 REP	W	99LSD047	09/15/99	09/19/99	09/20/99
SULFIDE	001 MS	W	99LSD047	09/15/99	09/19/99	09/20/99
SULFIDE	001 MSD	W	99LSD047	09/15/99	09/19/99	09/20/99
LAB QC:						
CHLORIDE BY IC	MB1	W	99LICB79	N/A	09/23/99	09/23/99
CHLORIDE BY IC	MB1 BS	W	99LICB79	N/A	09/23/99	09/23/99

Recra LabNet - Lionville Laboratory
INORGANIC ANALYTICAL DATA PACKAGE FOR
TNU-HANFORD B99-085

DATE RECEIVED: 09/17/99 RFW LOT # :9909L126

CLIENT ID /ANALYSIS	RFW #	MTX	PREP #	COLLECTION	EXTR/PREP	ANALYSIS
FLUORIDE BY IC	MB1	W	99LICB79	N/A	09/23/99	09/23/99
FLUORIDE BY IC	MB1 BS	W	99LICB79	N/A	09/23/99	09/23/99
NITRITE BY IC	MB1	W	99LICB79	N/A	09/23/99	09/23/99
NITRITE BY IC	MB1 BS	W	99LICB79	N/A	09/23/99	09/23/99
NITRATE BY IC	MB1	W	99LICB79	N/A	09/23/99	09/23/99
NITRATE BY IC	MB1 BS	W	99LICB79	N/A	09/23/99	09/23/99
PHOSPHATE BY IC	MB1	W	99LICC79	N/A	09/23/99	09/23/99
PHOSPHATE BY IC	MB1 BS	W	99LICC79	N/A	09/23/99	09/23/99
SULFATE BY IC	MB1	W	99LICB79	N/A	09/23/99	09/23/99
SULFATE BY IC	MB1 BS	W	99LICB79	N/A	09/23/99	09/23/99
NITRATE NITRITE	MB1	W	99LN3A47	N/A	10/01/99	10/01/99
NITRATE NITRITE	MB1 BS	W	99LN3A47	N/A	10/01/99	10/01/99
NITRATE NITRITE	MB1 BSD	W	99LN3A47	N/A	10/01/99	10/01/99
AMMONIA	MB1	W	99LAMA36	N/A	09/24/99	09/24/99
AMMONIA	MB1 BS	W	99LAMA36	N/A	09/24/99	09/24/99
AMMONIA	MB1 BSD	W	99LAMA36	N/A	09/24/99	09/24/99
SULFIDE	MB1	W	99LSD047	N/A	09/19/99	09/20/99
SULFIDE	MB1 BS	W	99LSD047	N/A	09/19/99	09/20/99
SULFIDE	MB1 BSD	W	99LSD047	N/A	09/19/99	09/20/99

RECRA LabNet Use Only
9909L126

Custody Transfer Record/Lab Work Request

Page 1 of 1

A11

FIELD PERSONNEL: COMPLETE ONLY SHADED AREAS

(8) perrone wet chem

RECRA LabNet

Client: TRU Hanford B99-085
Est. Final Proj. Sampling Date:
Project #: 10985-001-001-9999-00
Project Contact/Phone #:
RECRA Project Manager: OJ
QC: spec Del: std TAT: 30 day
Date Rec'd: 9-17-99 Date Due: 10/17/99
Account #:

	VOA	BNA	Pest/PCB	Herb				Metal	CN	Sulfide	IC Anions PH	NO2 NO3
Refrigerator #	1	6						6		6	6	6
#/Type Container Liquid	3V	2AG						3P		1P	1P	1P
#/Type Container Solid												
Volume Liquid	40ml	1L						1L		1L	1L	1L
Volume Solid										ZNAC		
Preservatives								HNO3		NaOH	–	H2SO4
RECRA LabNet Use Only	0024H OGCSC	0025H 0024H	9/17/99					Met①		1SFD	Ang①	1N3N2

ORGANIC: VOA, BNA, Pest/PCB, Herb. INORG: Metal, CN.

MATRIX CODES:
S - Soil
SE - Sediment
SO - Solid
SL - Sludge
W - Water
O - Oil
A - Air
DS - Drum Solids
DL - Drum Liquids
L - EP/TCLP Leachate
WI - Wipe
X - Other
F - Fish

Lab ID	Client ID/Description	MS	MSD	Matrix	Date Collected	Time Collected	VOA	BNA	Pest/PCB	Herb	Metal	CN	Sulfide	IC Anions	NO2 NO3
001	BOWCP8			W	9-15-99	0650	✓	✓			✓		✓	✓	✓
002	BOWCP9			↓	↓	0518	✓								

Special Instructions:
saf # B99-085
9/23/99 - INH3N added to 001 per client coc.

COMPOSITE WASTE

DATE/REVISIONS:
OGCSC = 1-propanol, Ethanol
Met①2. = as, Ba, Cd, Cr, Pb, Se, Ag, Cu,
3. Ni, V, Zn, Be
Ang①4. = ICCL, ICFL, ICNO2, ICNO3, ICPO4,
5. ICSO4, IPH, INH3N
6. Run matrix QC

Relinquished by	Received by	Date	Time
Fed Ex	TMurray	9-17-99	1020

Relinquished by	Received by	Date	Time
	ORIGINAL		
	REWRITTEN		

RECRA LabNet Use Only

Samples were:
1) Shipped ✓ or Hand Delivered ___
Airbill # *
2) Ambient or (Chilled)
3) Received in Good Condition (Y) or N
4) Labels Indicate Properly Preserved (Y) or N
5) Received Within Holding Times (IC, PH autofail) Y or (N)

COC Tape was:
1) Present on Outer Package (Y) or N
2) Unbroken on Outer Package (Y) or N
3) Present on Sample (Y) or N
4) Unbroken on Sample (Y) or N
COC Record Present Upon Sample Rec't (Y) or N
Cooler Temp. 4.1 °C

Discrepancies Between Samples Labels and COC Record? (Y) or N
NOTES: Sulfide bottle 1L not 500mL as chain indicates

*4235795295601

Bechtel Hanford Inc.		CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST	B99-085-04	Page 1 of 1

Collector Bowers/Trice	**Company Contact** C Cearlock **Telephone No.** 372-9574	**Project Coordinator** TRENT, SJ	**Price Code** 7N **Data Turnaround** 45 Days
Project Designation 200 Area Source characterization - 200-CW-1 OU - QC Sa	**Sampling Location** 200 East	**SAF No.** B99-085	
Ice Chest No. ERC 96 024	**Field Logbook No.** FCL511	**Method of Shipment** Fed Ex	
Shipped To TMA/RECRA B70 9-15-99	**Offsite Property No.** A990259	**Bill of Lading/Air Bill No.** 4285795295 61	
		COA B20CW1 671C	

POSSIBLE SAMPLE HAZARDS/REMARKS

Special Handling and/or Storage

Preservation	ZnAc+NaOH to pH >9 Cool	Cool 4C	H2SO4 to pH <2 Cool 4C	Cool 4C	HNO3 to pH <2	HCl to pH <2 Cool 4C	HNO3 to pH <2			
Type of Container	P	P	P	aG	P	aGs*	P			
No. of Container(s)	1	1	1	2	2	3	3			
Volume	500mL	1000mL	1000mL	1000mL	1000mL	40mL	500mL			
SAMPLE ANALYSIS	Sulfides - 9030	See item (1) in Special Instructions	NO2/NO3 - 353.1; Ammonia - 350.3	Semi-VOA - 8270A (TCL)	Gross Alpha, Gross Beta	VOA - 8260A (TCL); VOA - 8260A (Add-On) {1-Propanol, Ethanol}	See item (2) in Special Instructions.			

Sample No.	Matrix *	Sample Date	Sample Time										
B0WCP8	Water	9.15.99	0650	X	X	X	X		X	X			
B0WCP9	Water	9.15.99	0518						X				

CHAIN OF POSSESSION — **Sign/Print Names**

Relinquished By / Date/Time	Received By / Date/Time
Doug Bowers 9-17-99/1530	Ref 1B 9-17-99/1730
Ref 1B 91699 1300	SJ Gale SJ Gale 91699 1300
SJ Gale SJ Gale 91699 1300	FED EX
FedEx 9-17-99 1020	TMurray 9-17-99 10:20

SPECIAL INSTRUCTIONS

(1) IC Anions - 300.0 {Chloride, Fluoride, Nitrate, Nitrite, Phosphate, Sulfate}; pH (Water) - 9040
(2) ICP Metals - 6010A (Supertrace) {Arsenic, Barium, Cadmium, Chromium, Lead, Selenium, Silver}; ICP Metals - 6010A (Supertrace Add-On) {Copper, Nickel, Vanadium, Zinc} Be

samples from non rad area

COLLECTOR UNAVAILABLE TO SIGN COC.

Matrix *
Soil
Water
Vapor
Other Solid
Other Liquid

LABORATORY SECTION	Received By	Title	Date/Time
FINAL SAMPLE DISPOSITION	Disposal Method	Disposed By	Date/Time

012

Case Narrative

1.0 GENERAL

Bechtel Hanford Inc. Sample Delivery Group H0535 is composed of one liquid (water) sample designated under SAF No. B99-085 with a Project Designation of: 200 Area Source characterization-200-CW-1 OU-QC Sa.

The sample was received as stated on the Chain-of-Custody document. Any discrepancies are noted on the TNU Sample Receipt Checklist. The results were transmitted to BHI via facsimile on October 28, 1999.

2.0 ANALYSIS NOTES

2.1 Gross Alpha and Beta Analyses

No problems were encountered during the course of the analyses.



TMA/RICHMOND

SAMPLE DELIVERY GROUP H0535

SAMPLE SUMMARY

SDG 7213
Contact Kevin C. Johnson

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

CLIENT SAMPLE ID	LOCATION	MATRIX	LEVEL	LAB SAMPLE ID	SAF NO	CHAIN OF CUSTODY	COLLECTED
B0WCP8	200 East	WATER		N909119-01	B99-085	B99-085-04	09/15/99 06:50
Method Blank		WATER		N909119-03	B99-085		
Lab Control Sample		WATER		N909119-02	B99-085		
Duplicate (N909119-01)	200 East	WATER		N909119-04	B99-085		09/15/99 06:50

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-CS
Version 3.06
Report date 10/28/99

TMA/RICHMOND

SAMPLE DELIVERY GROUP H0535

QC SUMMARY

SDG 7213
Contact Kevin C. Johnson

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

QC BATCH	CHAIN OF CUSTODY	CLIENT SAMPLE ID	MATRIX	% SOLIDS	SAMPLE AMOUNT	BASIS AMOUNT	DAYS SINCE RECEIVED	DAYS SINCE COLL	LAB SAMPLE ID	DEPARTMENT SAMPLE ID
7213	B99-085-04	B0WCP8	WATER				09/17/99	2	N909119-01	7213-001
		Method Blank	WATER						N909119-03	7213-003
		Lab Control Sample	WATER						N909119-02	7213-002
		Duplicate (N909119-01)	WATER				09/17/99	2	N909119-04	7213-004

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-QS
Version 3.06
Report date 10/28/99

TMA/RICHMOND

SAMPLE DELIVERY GROUP H0535

PREP BATCH SUMMARY

SDG 7213
Contact Kevin C. Johnson

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

TEST	MATRIX	METHOD	PREPARATION BATCH	ERROR 2σ %	PLANCHETS ANALYZED CLIENT	MORE	RE	BLANK	LCS	DUP/ORIG	MS/ORIG	QUALIFIERS
Gas Proportional Counting												
80A	WATER	Gross Alpha in Water	6904-016	20.0	1			1	1	1/1		
80B	WATER	Gross Beta in Water	6904-016	15.0	1			1	1	1/1		

Duplicates and Matrix Spikes are those with original (Client) sample in this Sample Delivery Group.
Blank and LCS planchets are those in the same preparation batch as some Client, Duplicate or Spike sample.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-PBS
Version 3.06
Report date 10/28/99

TMA/RICHMOND

SAMPLE DELIVERY GROUP H0535

WORK SUMMARY

SDG 7213
Contact Kevin C. Johnson

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

CLIENT SAMPLE ID / LOCATION / CUSTODY	SAF No	MATRIX	LAB SAMPLE ID / COLLECTED / RECEIVED	PLANCHET	TEST	SUF-FIX	ANALYZED	REVIEWED	BY	METHOD
B0WCP8			N909119-01	7213-001	80A/80		10/25/99	10/28/99	NJV	Gross Alpha in Water
200 East		WATER	09/15/99	7213-001	80B/80		10/25/99	10/28/99	NJV	Gross Beta in Water
B99-085-04	B99-085		09/17/99							
Method Blank			N909119-03	7213-003	80A/80		10/25/99	10/28/99	NJV	Gross Alpha in Water
		WATER		7213-003	80B/80		10/25/99	10/28/99	NJV	Gross Beta in Water
	B99-085									
Lab Control Sample			N909119-02	7213-002	80A/80		10/25/99	10/28/99	NJV	Gross Alpha in Water
		WATER		7213-002	80B/80		10/25/99	10/28/99	NJV	Gross Beta in Water
	B99-085									
Duplicate (N909119-01)			N909119-04	7213-004	80A/80		10/25/99	10/28/99	NJV	Gross Alpha in Water
200 East		WATER	09/15/99	7213-004	80B/80		10/25/99	10/28/99	NJV	Gross Beta in Water
	B99-085		09/17/99							

COUNTS OF TESTS BY SAMPLE TYPE

TEST	SAF No	METHOD	REFERENCE	CLIENT	MORE	RE	BLANK	LCS	DUP SPIKE	TOTAL
80A/80	B99-085	Gross Alpha in Water	EPA900.0	1			1	1	1	4
80B/80	B99-085	Gross Beta in Water	EPA900.0	1			1	1	1	4
TOTALS				2			2	2	2	8

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-CWS
Version 3.06
Report date 10/28/99

TMA / RICHMOND

SAMPLE DELIVERY GROUP H0535

N909119-03 Method Blank

METHOD BLANK

SDG 7213	Client/Case no Hanford	SDG H0535	
Contact Kevin C. Johnson	Contract TRB-SBB-207925		
Lab sample id N909119-03	Client sample id Method Blank		
Dept sample id 7213-003	Material/Matrix	WATER	
	SAF No B99-085		

ANALYTE	CAS NO	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI-FIERS	TEST
Gross Alpha	12587-46-1	-0.321	0.34	1.1	3.0	U	80A
Gross Beta	12587-47-2	-0.325	1.4	2.4	4.0	U	80B

200 Area Source chtztn-200CW-1OUQCSa

QC-BLANK 32078

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-DS
Version 3.06
Report date 10/28/99

TMA/RICHMOND

SAMPLE DELIVERY GROUP H0535

N909119-02 Lab Control Sample

LAB CONTROL SAMPLE

SDG 7213 Client/Case no Hanford SDG H0535

Contact Kevin C. Johnson Case no TRB-SBB-207925

Lab sample id N909119-02 Client sample id Lab Control Sample

Dept sample id 7213-002 Material/Matrix WATER

SAF No B99-085

ANALYTE	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI-FIERS	TEST	ADDED pCi/L	2σ ERR pCi/L	REC %	3σ LMTS (TOTAL)	PROTOCOL LIMITS
Gross Alpha	61.6	4.9	1.4	3.0		80A	72.0	2.9	86	72-128	80-120
Gross Beta	85.3	3.8	1.9	4.0		80B	83.0	3.3	103	75-125	

200 Area Source chtztn-200CW-1OUQCSa

QC-LCS 32077

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-LCS
Version 3.06
Report date 10/28/99

TMA/RICHMOND

SAMPLE DELIVERY GROUP H0535

N909119-04 B0WCP8

DUPLICATE

SDG 7213 Client/Case no Hanford SDG H0535

Contact Kevin C. Johnson Case no TRB-SBB-207925

DUPLICATE		ORIGINAL			
Lab sample id	N909119-04	Lab sample id	N909119-01	Client sample id	B0WCP8
Dept sample id	7213-004	Dept sample id	7213-001	Location/Matrix	200 East WATER
		Received	09/17/99	Collected	09/15/99 06:50
				Custody/SAF No	B99-085-04 B99-085

ANALYTE	DUPLICATE pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI-FIERS	TEST	ORIGINAL pCi/L	2σ ERR (COUNT)	MDA pCi/L	QUALI-FIERS	RPD %	3σ PROT TOT LIMIT
Gross Alpha	-0.067	0.31	0.76	3.0	U	80A	-0.123	0.33	0.81	U	-	
Gross Beta	-0.742	1.3	2.3	4.0	U	80B	-0.167	1.2	2.1	U	-	

200 Area Source chtztn-200CW-1OUQCSa

QC-DUP#1 32079

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-DUP
Version 3.06
Report date 10/28/99

TMA / RICHMOND

SAMPLE DELIVERY GROUP H0535

N909119-01 B0WCP8

DATA SHEET

SDG	7213	Client/Case no	Hanford SDG H0535
Contact	Kevin C. Johnson	Contract	TRB-SBB-207925
Lab sample id	N909119-01	Client sample id	B0WCP8
Dept sample id	7213-001	Location/Matrix	200 East WATER
Received	09/17/99	Collected	09/15/99 06:50
		Custody/SAF No	B99-085-04 B99-085

ANALYTE	CAS NO	RESULT pCi/L	2σ ERR (COUNT)	MDA pCi/L	RDL pCi/L	QUALI-FIERS	TEST
Gross Alpha	12587-46-1	-0.123	0.33	0.81	3.0	U	80A
Gross Beta	12587-47-2	-0.167	1.2	2.1	4.0	U	80B

200 Area Source chtztn-200CW-1OUQCSa

Lab id	TMANC
Protocol	Hanford
Version	Ver 1.0
Form	DVD-DS
Version	3.06
Report date	10/28/99

TMA/RICHMOND

SAMPLE DELIVERY GROUP H0535

METHOD SUMMARY

GROSS ALPHA IN WATER

GAS PROPORTIONAL COUNTING

Test 80A Matrix WATER
SDG 7213
Contact Kevin C. Johnson

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

RESULTS

CLIENT SAMPLE ID	LAB SAMPLE ID	RAW TEST	SUF-FIX	PLANCHET	Gross Alpha	
Preparation batch 6904-016						
B0WCP8	N909119-01	80		7213-001	U	
BLK (QC ID=32078)	N909119-03	80		7213-003	U	
LCS (QC ID=32077)	N909119-02	80		7213-002	ok	
Duplicate (N909119-01)	N909119-04	80		7213-004	-	U
Nominal values and limits from method				RDLs (pCi/L)	3.0	
200 Area Source chtztn-200CW-1OUQCSa						

METHOD PERFORMANCE

CLIENT SAMPLE ID	LAB SAMPLE ID	RAW TEST	SUF-FIX	MDA pCi/L	ALIQ L	PREP FAC	DILU-TION	RESID mg	EFF %	COUNT min	FWHM keV	DRIFT KeV	DAYS HELD	PREPARED	ANAL-YZED	DETECTOR
Preparation batch 6904-016 2σ prep error 20.0 % Reference Lab Notebook 6904 pg. 016																
B0WCP8	N909119-01	80		0.81	0.300			1		100			40	10/20/99	10/25	GRB-114
BLK (QC ID=32078)	N909119-03	80		1.1	0.300			32		100				10/20/99	10/25	GRB-116
LCS (QC ID=32077)	N909119-02	80		1.4	0.300			36		100				10/20/99	10/25	GRB-115
Duplicate (N909119-01) (QC ID=32079)	N909119-04	80		0.76	0.300			2		100			40	10/20/99	10/25	GRB-112
Nominal values and limits from method				3.0	0.300			5-150		100			180			

PROCEDURES REFERENCE EPA900.0
EP-120 Gross Alpha and Gross Beta in Environmental Water, rev 2

AVERAGES ± 2 SD FOR 4 SAMPLES
MDA 1.0 ± 0.59
RESIDUE 18 ± 38

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-CMS
Version 3.06
Report date 10/28/99

TMA/RICHMOND

SAMPLE DELIVERY GROUP H0535

METHOD SUMMARY

GROSS BETA IN WATER

GAS PROPORTIONAL COUNTING

Test 80B Matrix WATER
SDG 7213
Contact Kevin C. Johnson

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

RESULTS

CLIENT SAMPLE ID	LAB SAMPLE ID	RAW TEST	SUF-FIX	PLANCHET	Gross Beta	
Preparation batch 6904-016						
B0WCP8	N909119-01	80		7213-001	U	
BLK (QC ID=32078)	N909119-03	80		7213-003	U	
LCS (QC ID=32077)	N909119-02	80		7213-002	ok	
Duplicate (N909119-01)	N909119-04	80		7213-004	-	U
Nominal values and limits from method				RDLs (pCi/L)	4.0	

200 Area Source chtztn-200CW-1OUQCSa

METHOD PERFORMANCE

CLIENT SAMPLE ID	LAB SAMPLE ID	RAW TEST	SUF-FIX	MDA pCi/L	ALIQ L	PREP FAC	DILU-TION	RESID mg	EFF %	COUNT min	FWHM keV	DRIFT KeV	DAYS HELD	PREPARED	ANAL-YZED	DETECTOR
Preparation batch 6904-016	2σ prep error 15.0 %	Reference Lab Notebook 6904 pg. 016														
B0WCP8	N909119-01	80		2.1	0.300			1		100			40	10/20/99	10/25	GRB-114
BLK (QC ID=32078)	N909119-03	80		2.4	0.300			32		100				10/20/99	10/25	GRB-116
LCS (QC ID=32077)	N909119-02	80		1.9	0.300			36		100				10/20/99	10/25	GRB-115
Duplicate (N909119-01) (QC ID=32079)	N909119-04	80		2.3	0.300			2		100			40	10/20/99	10/25	GRB-112
Nominal values and limits from method				4.0	0.300			5-150		100			180			

PROCEDURES REFERENCE EPA900.0
EP-120 Gross Alpha and Gross Beta in Environmental Water, rev 2

AVERAGES ± 2 SD FOR 4 SAMPLES
MDA 2.2 ± 0.44
RESIDUE 18 ± 38

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-CMS
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

REPORT GUIDE

Client Hanford
Contract TRB-SBB-207925
Case no SDG_H0535

SAMPLE SUMMARY

The Sample and QC Summary Reports show all samples, including QC samples, reported in one Sample Delivery Group (SDG).

The Sample Summary Report fully identifies client samples and gives the corresponding lab sample identification. The QC Summary Report shows at the sample level how the lab organized the samples into batches and generated QC samples. The Preparation Batch and Method Summary Reports show this at the analysis level.

The following notes apply to these reports:

* LAB SAMPLE ID is the lab's primary identification for a sample.

* DEPARTMENT SAMPLE ID is an alternate lab id, for example one assigned by a radiochemistry department in a lab.

* CLIENT SAMPLE ID is the client's primary identification for a sample. It includes any sample preparation done by the client that is necessary to identify the sample.

* QC BATCH is a lab assigned code that groups samples to be processed and QCed together. These samples should have similar matrices.

  QC BATCH is not necessarily the same as SDG, which reflects samples received and reported together.

* All Lab Control Samples, Method Blanks, Duplicates and Matrix Spikes are shown that QC any of the samples. Due to possible reanalyses, not all results for all these QC samples may be relevant to the SDG. The Lab Control Sample, Method Blank, Duplicate, Matrix Spike and Method Summary Reports detail these relationships.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

REPORT GUIDE

Client Hanford
Contract TRB-SBB-207925
Case no SDG_H0535

PREPARATION BATCH SUMMARY

The Preparation Batch Summary Report shows all preparation batches in one Sample Delivery Group (SDG) with information necessary to check the completeness and consistency of the SDG.

The following notes apply to this report:

* The preparation batches are shown in the same order as the Method Summary Reports are printed.

* Only analyses of planchets relevant to the SDG are included.

* Each preparation batch should have at least one Method Blank and LCS in it to validate client sample results.

* The QUALIFIERS shown are all qualifiers other than U, J, B, L and H that occur on any analysis in the preparation batch. The Method Summary Report has these qualifiers on a per sample basis.

  These qualifiers should be reviewed as follows:

  X Some data has been manually entered or modified. Transcription errors are possible.

  P One or more results are 'preliminary'. The data is not ready for final reporting.

  2 There were two or more results for one analyte on one planchet imported at one time. The results in DVD may not be the same as on the raw data sheets.

  Other lab defined qualifiers may occur. In general, these should be addressed in the SDG narrative.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

REPORT GUIDE

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

WORK SUMMARY

The Work Summary Report shows all samples, including QC samples, and all relevant analyses in one Sample Delivery Group (SDG). This report is often useful as supporting documentation for an invoice.

The following notes apply to this report:

* TEST is a code for the method used to measure associated analytes. Results and related information for each analyte are on the Data Sheet Report. In special cases, a test code used in the summary data section is not the same as in associated raw data. In this case, both codes are shown on the Work Summary.

* SUFFIX is the lab's code to distinguish multiple analyses (recounts, reworks, reanalyses) of a fraction of the sample. The suffix indicates which result is being reported. An empty suffix normally identifies the first attempt to analyze the sample.

* The LAB SAMPLE ID, TEST and SUFFIX uniquely identify all supporting data for a result. The Method Summary Report for each TEST has method performance data, such as yield, for each lab sample id and suffix and procedures used in the method.

* PLANCHET is an alternate lab identifier for work done for one test. It, combined with the TEST and SUFFIX, may be the best link to raw data.

* For QC samples, only analyses that directly QC some regular sample are shown. The Lab Control Sample, Method Blank, Duplicate, Matrix Spike and Method Summary Reports detail these relationships.

* The SAS (Special Analytical Services) Number is a client or lab assigned code that reflects special processing for samples, such as rapid turn around. Counts of tests done are lists by SAS number since it is likely to affect prices.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

REPORT GUIDE

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

DATA SHEET

The Data Sheet Report shows all results and primary supporting information for one client sample or Method Blank. This report corresponds to both the CLP Inorganics and Organics Data Sheet.

The following notes apply to this report:

* TEST is a code for the method used to measure an analyte. If the TEST is empty, no data is available; the analyte was not analyzed for.

* The LAB SAMPLE ID and TEST uniquely identify work within the Summary Data Section of a Data Package. The Work Summary and Method Summary Reports further identify raw data that underlies this work.

  The Method Summary Report for each TEST has method performance data, such as yield, for each Lab Sample ID and a list of procedures used in the method.

* ERRORs can be labeled TOTAL or COUNT. TOTAL implies a preparation (non-counting method) error has been added, as square root of sum of squares, to the counting error denoted by COUNT. The preparation errors, which may vary by preparation batch, are shown on the Method Summary Report.

* A RESULT can be 'N.R.' (Not Reported). This means the lab did this work but chooses not to report it now, possibly because it was reported at another time.

* When reporting a Method Blank, a RESULT can be 'N.A.' (Not Applicable). This means there is no reported client sample work in the same preparation batch as the Blank's result. This is likely to occur when the Method Blank is associated with reanalyses of selected work for a few samples in the SDG.

The following qualifiers are defined by the DVD system:

U The RESULT is less than the MDA (Minimum Detectable Activity).

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

GUIDE, cont.

Client Hanford
Contract TRB-SBB-207925
Case no SDG_H0535

DATA SHEET

If the MDA is blank, the ERROR is used as the limit.

J The RESULT is less than the RDL (Required Detection Limit) and no U qualifier is assigned.

B A Method Blank associated with this sample had a result without a U flag and, after correcting for possibly different aliquots, that result is greater than or equal to the MDA for this sample.

Normally, B is not assigned if U is. When method blank subtraction is shown on this report, B flags are assigned based on the unsubtracted values while U's are assigned based on the subtracted ones. Both flags can be assigned in this case.

For each sample result, all Method Blank results in the same preparation batch are compared. The Method Summary Report documents this and other QC relationships.

L Some Lab Control Sample that QC's this sample had a low recovery. The lab can disable assignment of this qualifier.

H Similar to 'L' except the recovery was high.

P The RESULT is 'preliminary'.

X Some data necessary to compute the RESULT, ERROR or MDA was manually entered or modified.

2 There were two or more results available for this analyte. The reported result may not be the same as in the raw data.

Other qualifiers are lab defined. Definitions should be in the SDG narrative.

The following values are underlined to indicate possible problems:

* An MDA is underlined if it is bigger than its RDL.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

GUIDE, cont.

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

DATA SHEET

* An ERROR is underlined if the 1.645 sigma counting error is bigger than both the MDA and the RESULT, implying that the MDA may not be a good estimate of the 'real' minimum detectable activity.

* A negative RESULT is underlined if it is less than the negative of its 2 sigma counting ERROR.

* When reporting a Method Blank, a RESULT is underlined if greater than its MDA. If the MDA is blank, the 2 sigma counting error is used in the comparison.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

REPORT GUIDE

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

LAB CONTROL SAMPLE

The Lab Control Sample Report shows all results, recoveries and primary supporting information for one Lab Control Sample.

The following notes apply to this report:

* All fields in common with the Data Sheet Report have similar usage. Refer to its Report Guide for details.

* An amount ADDED is the lab's value for the actual amount spiked into this sample with its ERROR an estimate of the error of this amount.

  An amount added is underlined if its ratio to the corresponding RDL is outside protocol specified limits.

* REC (Recovery) is RESULT divided by ADDED expressed as a percent.

* The first, computed limits for the recovery reflect:

  1. The error of RESULT, including that introduced by rounding the result prior to printing.

     If the limits are labeled (TOTAL), they include preparation error in the result. If labeled (COUNT), they do not.

  2. The error of ADDED.

  3. A lab specified, per analyte bias. The bias changes the center of the computed limits.

* The second limits are protocol defined upper and lower QC limits for the recovery.

* The recovery is underlined if it is outside either of these ranges.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

REPORT GUIDE

Client Hanford
Contract TRB-SBB-207925
Case no SDG_H0535

DUPLICATE

The Duplicate Report shows all results, differences and primary supporting information for one Duplicate and associated Original sample.

The following notes apply to this report:

* All fields in common with the Data Sheet Report have similar usage. This applies both to the Duplicate and Original sample data. Refer to the Data Sheet Report Guide for details.

  If the Duplicate has data for a TEST and the lab did not do this test to the Original, the Original's RESULTs are underlined.

* The RPD (Relative Percent Difference) is the absolute value of the difference of the RESULTs divided by their average expressed as a percent.

  If both RESULTs are less than their MDAs, no RPD is computed and a '-' is printed.

  For an analyte, if the lab did work for both samples but has data for only one, the MDA from the sample with data is used as the other's result in the RPD.

* The first, computed limit is the sum, as square root of sum of squares, of the errors of the results divided by the average result as a percent, hence the relative error of the difference rather than the error of the relative difference. The errors include those introduced by rounding the RESULTs prior to printing.

  If this limit is labeled TOT, it includes the preparation error in the RESULTs. If labeled CNT, it does not.

  This value reported for this limit is at most 999.

* The second limit for the RPD is the larger of:

  1. A fixed percentage specified in the protocol.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

GUIDE, cont.

Client Hanford
Contract TRB-SBB-207925
Case no SDG_H0535

DUPLICATE

2. A protocol factor (typically 2) times the average MDA as a percent of the average result. This limit applies when the results are close to the MDAs.

* The RPD is underlined if it is greater than either limit.

* If specified by the lab, the second limit column is replaced by the Difference Error Ratio (DER), which is the absolute value of the difference of the results divided by the quadratic sum of their one sigma errors, the same errors as used in the first limit.

  Except for differences due to rounding, the DER is the same as the RPD divided by the first RPD limit with the limit scaled to 1 sigma.

* The DER is underlined if it is greater than the sigma factor, typically 2 or 3, shown in the header for the first RPD limit.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

REPORT GUIDE

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

MATRIX SPIKE

The Matrix Spike Report shows all results, recoveries and primary supporting information for one Matrix Spike and associated Original sample.

The following notes apply to this report:

* All fields in common with the Data Sheet Report have similar usage. This applies both to the Spiked and Original sample data. Refer to the Data Sheet Report Guide for details.

  If the Spike has data for a TEST and the lab did not do this test to the Original, the Original's RESULTs are underlined.

* An amount ADDED is the lab's value for the actual amount spiked into the Spike sample with its ERROR an estimate of the error of this amount.

  An amount is underlined if its ratio to the corresponding RDL is outside protocol specified limits.

* REC (Recovery) is the Spike RESULT minus the Original RESULT divided by ADDED expressed as a percent.

* The first, computed limits for the recovery reflect:

  1. The errors of the two RESULTs, including those introduced by rounding them prior to printing.

     If the limits are labeled (TOTAL), they include preparation error in the result. If labeled (COUNT), they do not.

  2. The error of ADDED.

  3. A lab specified, per analyte bias. The bias changes the center of the computed limits.

* The second limits are protocol defined upper and lower QC limits

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

GUIDE, cont.

Client Hanford
Contract TRB-SBB-207925
Case no SDG_H0535

MATRIX SPIKE

for the recovery.

These limits are left blank if the Original RESULT is more than a protocol defined factor (typically 4) times ADDED. This is a way of accounting for that when the spike is small compared to the amount in the original sample, the recovery is unreliable.

* The recovery is underlined (out of spec) if it is outside either of these ranges.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

REPORT GUIDE

Client Hanford
Contract TRB-SBB-207925
Case no SDG_H0535

METHOD SUMMARY

The Method Summary Report has two tables. One shows up to five results measured using one method. The other has performance data for the method. There is one report for each TEST, as used on the Data Sheet Report.

The following notes apply to this report:

* Each table is subdivided into sections, one for each preparation batch. A preparation batch is a group of aliquots prepared at roughly the same time in one work area of the lab using the same method.

  There should be Lab Control Sample and Method Blank results in each preparation batch since this close correspondence makes the QC meaningful. Depending on lab policy, Duplicates need not occur in each batch since they QC sample dependencies such as matrix effects.

* The RAW TEST column shows the test code used in the raw data to identify a particular analysis if it is different than the test code in the header of the report. This occurs in special cases due to method specific details about how the lab labels work.

  The Lab Sample or Planchet ID combined with the (Raw) Test Code and Suffix uniquely identify the raw data for each analysis.

* If a result is less than both its MDA and RDL, it is replaced by just 'U' on this report. If it is greater than or equal to the RDL but less than the MDA, the result is shown with a 'U' flag.

  The J and X flags are as on the data sheet.

* Non-U results for Method Blanks are underlined to indicate possible contamination of other samples in the preparation batch. The Method Blank Report has supporting data.

* Lab Control Sample and Matrix Spike results are shown as: ok, No data, LOW or HIGH, with the last two underlined. 'No data'

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

GUIDE, cont.

Client Hanford
Contract TRB-SBB-207925
Case no SDG H0535

METHOD SUMMARY

means no amount ADDED was specified. 'LOW' and 'HIGH' correspond to when the recovery is underlined on the Lab Control Sample or Matrix Spike Report. See these reports for supporting data.

* Duplicate sample results are shown as: ok, No data, or OUT, with the last two underlined. 'No data' means there was no original sample data found for this duplicate. 'OUT' corresponds to when the RPD is underlined on the Duplicate Report. See this report for supporting data.

* If the MDA column is labeled 'MAX MDA', there was more than one result measured by the reported method and the MDA shown is the largest MDA. If not all these results have the same RDL, the MAX MDA reflects only those results with RDL equal to the smallest one.

  MDAs are underlined if greater than the printed RDL.

* Aliquots are underlined if less than the nominal value specified for the method.

* Prepareation factors are underlined if greater than the nominal value specified for the method.

* Dilution factors are underlined if greater than the nominal value specified for the method.

* Residues are underlined if outside the range specified for the method. Residues are not printed if yields are.

* Yields, which may be gravimetric, radiometric or some type of recovery depending on the method, are underlined if outside the range specified for the method.

* Efficiencies are underlined if outside the range specified for the method. Efficiencies are detector and geometry dependent so this test is only approximate.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

GUIDE, cont.

Client Hanford
Contract TRB-SBB-207925
Case no SDG_H0535

METHOD SUMMARY

* Count times are underlined if less than the nominal value specified for the method.

* Resolutions (as FWHM; Full Width at Half Max) are underlined if greater than the method specified limit.

* Tracer drifts are underlined if their absolute values are greater than the method specified limit. Tracer drifts are not printed if percent moistures are.

* Days Held are underlined if greater than the holding time specified in the protocol.

* Analysis dates are underlined if before their planchet's preparation date or, if a limit is specified, too far after it.

For some methods, ratios as percentages and error estimates for them are computed for pairs of results. A ratio column header like '1÷3' means the ratio of the first result column and the third result column.

Ratios are not computed for Lab Control Sample, Method Blank or Matrix Spike results since their matrices are not necessarily similar to client samples'.

The error estimate for a ratio of results from one planchet reflects only counting errors since other errors should be correlated. For a ratio involving different planchets, if QC limits are computed based on total errors, the error for the ratio allows for the preparation errors for the planchets.

The ratio is underlined (out of spec) if the absolute value of its difference from the nominal value is greater than its error estimate. If no nominal value is specified, this test is not done.

For Gross Alpha or Gross Beta results, there may be a column showing the sum of other Alpha or Beta emitters. This sum includes all relevant

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

SDG 7213
Contact Kevin C. Johnson

GUIDE, cont.

Client Hanford
Contract TRB-SBB-207925
Case no SDG_H0535

METHOD SUMMARY

results in the DVD database, whether reported or not. Results in the sum are weighted by a particles/decay value specified by the lab for each relevant analyte. Results less than their MDA are not included. No sums are computed for Lab Control, Method Blank or Matrix Spike samples since their various planchets may not be physically related.

If a ratio of total isotopic to Gross Alpha or Beta is shown, the error for the ratio reflects both the error in the Gross result and the sum, as square root of sum of squares, of the errors in the isotopic results.

For total elemental uranium or thorium results, there may be a column showing the total weight computed from associated isotopic results. Ignoring results less than their MDAs, this is a weighted sum of the isotopic results. The weights depend on the molecular weight and half-life of each isotope so as to convert activities (decays) to weight (atoms).

If a ratio of total computed to measured elemental uranium or thorium is shown, the error for the ratio reflects the errors in all the measurements.

Lab id TMANC
Protocol Hanford
Version Ver 1.0
Form DVD-RG
Version 3.06
Report date 10/28/99

Bechtel Hanford Inc.	CHAIN OF CUSTODY/SAMPLE ANALYSIS REQUEST	B99-085-04	Page 1 of 1

Collector: Bowers/Trice	Company Contact: C. Clearlock	Telephone No: 372-9574	Project Coordinator: TRENT, SJ
Project Designation: 200 Area Source characterization - 200-CW-1 OU - QC Sa	Sampling Location: 200 East	SAF No.: B99-085	Price Code: 7N
Ice Chest No.: ERC 96 087	Field Logbook No.: EL 1511	Method of Shipment: Fed Ex	Data Turnaround: 45 Days
Shipped To: TMA/RECRA 9-15-99	Offsite Property No.: A99 0258	Bill of Lading/Air Bill No.: 423579529539	
		COA: B20CW1 67IC	

POSSIBLE SAMPLE HAZARDS/REMARKS

Special Handling and/or Storage

Sample No.	Matrix *	Sample Date	Sample Time	Preservation	Type of Container	No. of Container(s) / Volume	SAMPLE ANALYSIS								
				ZnAc+NaOH to pH >9 Cool	P	1 500mL	Sulfides - 9030								
				Cool 4C	P	1 1000mL	See item (1) in Special Instructions.								
				H2SO4 to pH <2 Cool 4C	P	1 1000mL	NO2/NO3 - 353.1; Ammonia - 350.3								
				Cool 4C	aG	2 1000mL	Semi-VOA - 8270A (TCL)								
				HNO3 to pH <2	P	2 1000mL	Gross Alpha; Gross Beta								
				HCl to pH <2 Cool 4C	aGs*	3 40mL	VOA - 8260A (TCL); VOA - 8260A (Add-On) (1-Propanol, Ethanol)								
				HNO3 to pH <2	P	3 500mL	See item (2) in Special Instructions.								
B0WCP8	Water	9.15.99	0650								X				
~~B0WCP9~~ SAS 9/16/99	Water	9.15.99	0518												

CHAIN OF POSSESSION

Sign/Print Names		
Relinquished By: Dave Bowers / Dave Bowers	Date/Time: 9-15-99/1530	Received By: B. F. [illegible]	Date/Time: 9-15-99/1530
Relinquished By: Kaf I-B	Date/Time: 9.16.99 1200	Received By: R. Fahlberg / R. Fahlberg	Date/Time: 9.16.99 1200
Relinquished By: R. Fahlberg / R. Fahlberg	Date/Time: 1430 9.16.99	Received By: Fed Ex	Date/Time: 9-16-99
Relinquished By: FedEx	Date/Time: 9-17-99 11:00	Received By: TNV M. Goldenberg	Date/Time: 11:00 9-17-99

SPECIAL INSTRUCTIONS

(1) IC Anions - 300.0 (Chloride, Fluoride, Nitrate, Nitrite, Phosphate, Sulfate); pH (Water) - 9040
(2) ICP Metals - 6010A (Supertrace) (Arsenic, Barium, Cadmium, Chromium, Lead, Selenium, Silver); ICP Metals - 6010A (Supertrace Add-On) (Copper, Nickel, Vanadium, Zinc)

samples from non prod area

collector unavailable to sign COC.

Matrix *
Soil
Water
Vapor
Other Solid
Other Liquid

LABORATORY SECTION	Received By	Title	Date/Time
FINAL SAMPLE DISPOSITION	Disposal Method	Disposed By	Date/Time

Thermo NUtech - Richmond

SAMPLE RECEIPT CHECKLIST

SAMPLE RECEIPT

Client: Bechtel Hanford Inc Date/Time received 9-17-99 11:00

CoC No. B99-085-04

Container I.D. No. ERC 96-087 Requested TAT (Days) 45 P.O. Received Yes [ ] No [✓]

INSPECTION

1.	Custody seals on shipping container intact?	Yes [✓]	No [ ]	N/A [ ]
2.	Custody seals on shipping container dated & signed?	Yes [✓]	No [ ]	N/A [ ]
3.	Custody seals on sample containers intact?	Yes [✓]	No [ ]	N/A [ ]
4.	Custody seals on sample containers dated & signed?	Yes [✓]	No [ ]	N/A [ ]
5.	Cooler Temperature: _____	Packing material is:	Wet [ ]	Dry [✓]
6.	Number of samples in shipping container: 1			
7.	Number of containers per sample: 2 (Or see CoC _____)			
8.	Paperwork agrees with samples?	Yes [✓]	No [ ]	
9.	Samples have: Tape [ ] Hazard labels [ ] Rad labels [ ] Appropriate sample labels [✓]			
10.	Samples are: In good condition [✓] Leaking [ ] Broken Container [ ] Missing [ ]			
11.	Describe any anomalies: _____			
13.	Was P.M. notified of any anomalies? Yes [ ] No [ ] Date _____			
14.	Received by M. Goldenberg Date: 9-17-99 Time: 11:00			

LOGIN

TNU W.O. No. _____ Group No. _____ Client W.O. No. _____

PROGRAM MANAGER

Sample holding times exceeded? Yes [ ] No [ ]

Client Notified: Name _____ Date/time _____